Development and Validation of a GC-FID Method for the Ouantitation of 20 Different Acidic and Neutral Cannabinoids

Authors

Waseem Gul¹, Elsayed A. Ibrahim^{2,3}, Sharjeel W. Gul^{1,4,5,6}, Iram Shahzadi¹, Mohamed M. Radwan², Suman Chandra², Hemant Lata², Mahmoud A. ElSohly^{1,2,5,7}

Affiliations

- 1 ElSohly Laboratories, Inc., 5 Industrial Park Drive, Oxford, MS, USA
- 2 National Center for Natural Products Research, School of Pharmacy, The University of Mississippi, University, MS, USA
- 3 Department of Pharmaceutical Analytical Chemistry, Faculty of Pharmacy, Suez Canal University, Ismailia, Egypt
- 4 Sally McDonnell Barksdale Honors College, University of Mississippi, USA
- 5 School of Pharmacy, University of Mississippi, USA
- 6 Frontier Medical College, Abbottabad, KPK, Pakistan
- 7 Department of Pharmaceutics and Drug Delivery, School of Pharmacy, The University of Mississippi, University, USA

Key words

Cannabis sativa, Cannabaceae, cannabinoids, GC-FID, derivatization

received	June 2, 2022
accepted after revision	October 18, 2022
published online	January 24, 2023

Bibliography

Planta Med 2023; 89: 683-696 DOI 10.1055/a-1962-8165 ISSN 0032-0943 © 2023. Thieme. All rights reserved. Georg Thieme Verlag KG, Rüdigerstraße 14, 70469 Stuttgart, Germany

Correspondence

Prof. Dr. Mahmoud A. ElSohly ElSohly Laboratories, Inc. 5 Industrial Park Drive, 38655 Oxford, MS, USA Phone: +16622362609, Fax: +16622344253 elsohly@elsohly.com



Supplementary material is available under https://doi.org/10.1055/a-1962-8165

ABSTRACT

For decades, Cannabis sativa had been illegal to sell or consume around the world, including in the United States. However, in light of the recent 2018 Farm Bill and the legalization of hemp across the US, various cannabis preparations have flooded the market, making it essential to be able to quantitate the levels of the different acidic and neutral cannabinoids in C. sativa and to have a complete cannabinoid profile of the different chemovars of the cannabis plant. A GC-FID method was developed and validated for the analysis of 20 acidic and neutral cannabinoids as trimethylsilyl (TMS) derivatives. The analyzed cannabinoids include cannabidivarinic acid (CBDVA), cannabidiolic acid (CBDA), cannabinolic acid (CBNA), cannabielsoic acid (CBEA), cannabicyclolic acid (CBLA), cannabichromenic acid (CBCA), trans-Δ⁹-tetrahydrocannabivarinic acid (Δ^9 -THCVA), trans- Δ^9 -tetrahydrocannabinolic acid A (Δ^9 -THCAA), cannabigerolic acid (CBGA), cannabidiol (CBD), cannabicyclol (CBL), cannabidivarin (CBDV), trans-Δ9-tetrahydrocannabivarin (THCV), cannabichromene (CBC), trans- Δ^{8} -tetrahydrocannabinol (Δ^{8} -THC), trans- Δ^{9} -tetrahydrocannabinol (Δ^9 -THC), cannabigerol (CBG), cannabinol (CBN), cannabicitran (CBT), and cannabielsoin (CBE). The method limit of detection (LOD) was as low as $0.1 \,\mu g/mL$, while the limit of quantitation ranged from 0.25 µg/mL to 0.5 µg/mL. The precision (%RSD) was <10%, while trueness ranged from 90-107%. The developed method is simple, accurate, and sensitive for the quantitation of all 20 acidic and neutral cannabinoids. Finally, the proposed method was successfully applied to the quantitation of the cannabinoids in different cannabis chemovars grown at the University of Mississippi.

Introduction

Cannabis (Cannabis sativa L., Cannabaceae) has been legalized in most states of the US, either for medical or recreational uses [1]. It is indicated to be used for the treatment of an enormous array of health problems including pain, inflammation, amenorrhea, and arthritis [2]. The extended history of medical use of cannabis in the treatment of many symptoms and diseases is attributed to its rich content of phytochemicals, namely, cannabinoids in addi-

ABBREVIATIONS

ADDICLVIA	
BSTFA	N,O-Bis (trimethylsilyl)-trifluoroacetamide
C. sativa	Cannabis sativa
CBC	cannabichromene
CBCA	cannabichromenic acid
CBD	cannabidiol
CBDA	cannabidiolic acid
CBDV	cannabidivarian
CBDVA	cannabidivarinic acid
CBE	cannabielsoin
CBEA	cannabielsoic acid
CBG	cannabigerol
CBGA	cannabigerolic acid
CBL	cannabicyclol
CBLA	cannabicyclolic acid
CBN	cannabinol
CBNA	cannabinolic acid
CBT	cannabicitran
DMAP	dimethylaminopyridine
THCV	$trans$ - Δ^9 -tetrahydrocannabivarian
TMS	trimethylsilyl
∆ ⁸ -THC	$trans$ - Δ^8 -tetrahydrocannabinol
∆ ⁹ -THC	$trans$ - Δ^9 -tetrahydrocannabinol
Δ ⁹ -THCAA	, , , , , , , , , , , , , , , , , , ,
Δ ⁹ -THCVA	$trans$ - Δ^9 -tetrahydrocannabivarinic acid

tion to the non-cannabinoids, such as terpenes and flavonoids, and to a lesser extent alkaloids [3].

The most important constituents of cannabis are the cannabinoids, such as Δ^9 -tetrahydrocannabinol (Δ^9 -THC), the main psychoactive cannabinoid currently approved to treat adverse effects associated with chemotherapy, in addition to the non-psychoactive component, cannabidiol (CBD), which is approved by FDA for the treatment of seizures [4, 5].

Cannabis has been classified into three chemovars. The first class is the high-THC chemovar that contains Δ^9 -THC as the predominant cannabinoid. The second chemovar is the CBD dominant, commonly known as hemp, with <0.3% THC level. The third chemovar is an intermediate chemovar with balanced levels of both THC and CBD with levels ≥ 1% of each [6].

Currently, hemp is extensively used in the manufacture of a variety of commercial products containing CBD. These include edibles, topicals, cosmetics, and hair products, as well as a variety of foods and drinks [7–9]. The US Food and Drug Administration (FDA) approved some cannabis-based medicines like dronabinol (Marinol) and nabilone (Cesamet) for the treatment of chemotherapy-induced nausea and vomiting in acquired immune deficiency syndrome (AIDS) patients [10, 11]. Also, Epidiolex[®] (CBDsolution) has been approved by the FDA for the treatment of Lennox-Gastaut and Dravet's syndrome, especially in children [12, 13]. According to the European Medicines Agency (EMA), there is only one cannabis-based medicine – Epidiolex[®] – which can be used in the European Union for the treatment of rare types of epilepsy (Lennox-Gastaut and Dravet's syndrome) and could also be used in the treatment of tuberous sclerosis [14].

Other cannabinoids have been shown to have important biological activities. For example, cannabigerol (CBG) has been tested in combination with THC and CBD as an anti-cancer agent and found to be a potent therapeutic agent in the treatment of glioblastoma [15]. Cannabidivarin (CBDV) was found to be an effective therapeutic agent in the treatment of Rett syndrome [16]. Also, Δ^9 -tetrahydrocannabivarin (THCV) was clinically reported to manage obesity and diabetes [17]. Cannabichromene (CBC) might be used as an alternative therapy in the treatment of acute respiratory distress syndrome (ARDS) [18]. In addition, CBC was reported as a cannabinoid CB2 receptor agonist and hence plays an important role in the modulation of inflammation [19].

Therefore, the chemical profile of cannabis, as a botanical drug, needs to be delineated and the reproducibility of the composition from batch to batch needs to be established by chemical analysis.

Searching the literature, many analytical methods were used for the chemical profiling of the cannabinoid content in cannabis. Some of these methods don't require derivatization before analysis, such as HPLC-UV and LC-MS, but might require more prepurification steps [20–28]. LC-MS/MS methods were used for the determination of illicit drugs in biological matrices [29]. In addition, LC-MS/MS was used for the analysis of most common narcotic substances in seized illicit materials and phytocannabinoids in oilbased preparations [30, 31].

While HPLC is recognized and used by many investigators as the primary method for analysis of cannabinoids because of its simplicity and ease of use, including no need for derivatization, the GC/FID outlined here offer a valid alternative with the advantage of increased sensitivity, peak resolution, and wide dynamic range [32].

GC-FID and GC-MS have been frequently used for the determination of cannabinoids [33, 34]. However, in general, GC techniques are unable to determine acids without prior derivatization, as decarboxylation occurs at the high temperature of the GC inlet [35–38].

To avoid decarboxylation and/or degradation of cannabinoids, chemical derivatization (silylation as TMS derivatives) is used. To the best of our knowledge, cannabinoids are most effectively identified and quantified using silylated derivatives [39, 40]. Derivatization enhances chromatography and repeatability by capturing polar or reactive groups in the target analyte. Furthermore, derivatization makes the analytes more volatile, therefore easier to be chromatographed at lower temperatures [36, 39, 41, 42].

To determine the intact acidic cannabinoids and avoid the decarboxylation process, samples have to be derivatized, most commonly by the silylation technique using N,O-bis(trimethylsilyl)trifluoroacetamide (BSTFA) [36, 43, 44].

We have previously published a GC-FID method for the analysis of 13 cannabinoids [45]. This study aimed to add another 7 cannabinoids to the previously published method, which enables the analysis of 9 acidic and 11 neutral cannabinoids, namely, CBDVA, THCVA, CBDA, CBCA, Δ^9 -THCAA, CBGA, CBNA, CBLA, CBEA, CBDV, THCV, CBD, CBC, Δ^8 -THC, Δ^9 -THC CBG, CBN, CBL, CBE,




Fig. 1 Chemical structures of the 20 silylated acidic and neutral cannabinoids

and CBT, and was applied for the analysis of plant material extracts.

Results and Discussion

As a continuation of our previous work, we thought it worthwhile to analyze more cannabinoids aiming at getting a better cannabinoid profile using GC-FID [45].

During the growth process of the *C. sativa* plant, phytocannabinoid acids are biosynthesized and, post-harvest, these acids decarboxylate to their neutral analogs when exposed to heat and light. In order to quantify acidic cannabinoids in addition to neutral ones, BSTFA was used as the silylating agent. The high derivatization/silylation rate of BSTFA for both neutral and acidic cannabinoids by reacting with both hydroxyl and carboxylic groups prevents acidic cannabinoids from decarboxylation at the high temperature of the injection port. In addition, silylation increases the method sensitivity and enhances peak symmetry. Dimethyl aminopyridine (DMAP) was used as a base or catalyst during the derivatization process. The chemical structures of the silylated cannabinoids are shown in **> Fig. 1**.

The method was proved to be sensitive, accurate, and fast as the 20 cannabinoid TMS derivatives are eluted in less than 11 min (with a total run of 17.5 min) without sacrificing resolution between adjacent peaks. A GC chromatogram of a mixture of the 20 cannabinoid TMS derivatives is shown in \triangleright Fig. 2a.

To determine the degree of carry-over, one blank TMS was injected immediately after each sample (Fig. 1S, Supporting Information). The blank TMS sample must not show the analytes and/ or internal standard peaks at a signal-to-noise (S/N) ratio of \geq 3. Fig. 2S (Supporting Information) shows the peak of IS at the expected retention time. The linear equation regression parameters are shown in **> Table 1**.

In order to confirm the applicability of this work, the developed method was then validated following ICH guidelines [46]. The sixpoint calibration curves were constructed, in triplicate, for each silylated cannabinoid using the internal standard (IS) method. Calibration curves were linear over the dynamic concentration range of 0.25–25 µg/mL for all silylated cannabinoids except for




Fig. 2 a Representative chromatogram of the standard silylated cannabinoid mixture (25 μg/mL) and IS (50 μg/mL) with respective retention times, **b** representative chromatogram of the high THC chemovar, **c** representative chromatogram of the intermediate chemovar, and **d** representative chromatogram of the high CBD chemovar.

CBE, CBEA, CBN, CBNA, and CBLA, where the dynamic range was $0.5-25 \mu g/mL$. It is obvious that all compounds had a coefficient of determination ($R^2 > 0.999$) indicating an excellent fit of the silylated cannabinoids to the model within the range studied (**> Table 1**).

The limits of detection (LOD) and quantitation (LOQ) are shown in \triangleright **Table 2**. The overall LOD and LOQ of the method were found to be 0.1 µg/mL and 0.25 µg/mL, respectively, for all canna-

binoids except the minors (CBN, CBLA, CBE, CBEA, and CBLA), where the LOQ was $0.5 \,\mu$ g/mL (**> Table 2**).

The intra-day trueness (evaluated as % recovery) and precision expressed as relative standard deviation (%RSD) ranged between 90–107% and 0.4–9.10%, respectively, for all cannabinoids. The inter-day trueness and precision ranged from 93–105% and 0.4–5.8%, respectively (**▶ Tables 3** and **4**). These values indicate that

Parameters	Calibration range (µg/mL)	Regression equation (Y): SLOPE (S)	Correlation coefficient (R ²)
CBDV	0.25–25	y = 0.0313 x - 0.0008	1.0000
THCV	0.25–25	y = 0.0270 x - 0.0002	1.0000
CBT	0.25–25	y = 0.0189 x + 0.0005	0.9999
CBD	0.25–25	y = 0.0317 x + 0.0002	0.9999
CBL	0.25–25	y = 0.0252 x + 0.0002	0.9999
CBC	0.25–25	y = 0.0305 x + 0.0022	0.9998
∆ ⁸ -THC	0.25–25	y = 0.0292 x + 0.0028	0.9997
∆ ⁹ -THC	0.25–25	y = 0.0277 x + 0.0031	0.9999
CBDVA	0.25–25	y = 0.0280 x - 0.0005	0.9998
CBN	0.5–25	y = 0.0277 x + 0.0006	0.9999
CBG	0.25–25	y = 0.0318 x - 0.0006	0.9999
CBE	0.5–25	y = 0.023 x + 0.0072	0.9995
THCVA	0.25–25	y = 0.0244 x - 0.0005	0.9999
CBDA	0.25–25	y = 0.0488 x - 0.0014	0.9993
CBLA	0.5–25	y = 0.007 x + 0.0013	0.9997
THCAA	0.25–10	y = 0.0187 x - 0.0015	0.9997
CBCA	0.25–25	y = 0.0212 x + 0.0000	0.9999
CBGA	0.25–25	y = 0.0268 x + 0.0008	0.9998
CBEA	0.5-0.25	y = 0.0126 x - 0.0004	0.9996
CBNA	0.5–25	y = 0.0250 x - 0.0008	1.000

this method is accurate, precise, and acceptable for the quantification of these cannabinoids.

The validated method was subsequently applied for the analysis of plant material from different cannabis chemovars (high-CBD chemovar, intermediate chemovar, and high-THC chemovar). The analysis results are shown in > Table 5. The results of the analysis of the three cannabis chemovars in terms of the content of the major cannabinoids (THC and/or CBD) were different. From the results, it is clear that the validated method was able to determine acidic cannabinoids in addition to neutral cannabinoids. The values of THCAA and Δ^9 -THC were higher in high-THC chemovar and found to be about 20% and 2%, respectively (> Table 5, and ▶ Fig. 2b). In high-THC chemovar, other cannabinoids found were CBG, THCVA, CBEA, and CBGA. The intermediate chemovar contained nearly equal amounts of CBD/CBDA and Δ^9 -THC/THCAA (> Fig. 2 c). The high-CBD chemovar contained high amounts of CBD/CBDA in comparison with other cannabinoids (> Fig. 2 d). The cannabinoids' content of the different chemovars is shown in > Table 5.

In conclusion, the developed GC-FID method is simple, accurate, and sensitive for the analysis of 20 acidic and neutral cannabinoids with baseline separation of all analytes. It can be routinely used in laboratories for quick and accurate analysis of cannabinoid content.

Material and Methods

Standards and reagents

Seventeen cannabinoids' reference standards were purchased from Cerilliant. Neutral cannabinoids (CBC, CBL, CBD, CBDV, CBG, CBN, THCV, Δ^8 -THC, and Δ^9 -THC) were at 1 mg/mL in MeOH, while the acidic cannabinoids (CBCA, CBDA, CBDVA, CBGA, CBLA, CBNA, Δ^9 -THCAA, and Δ^9 -THCVA) were at 1 mg/mL in MeCN. CBE, CBEA, and CBT were isolated from the cannabis plant material, identified, and confirmed by ¹H and ¹³C NMR (Coy Waller complex, National Center for Natural Products Research (NCNPR), School of Pharmacy, The University of Mississippi), which were prepared at 1 mg/mL in MeCN. The purity of all the reference standards was confirmed using GC-FID and GC-MS (purity >98%). The internal standard 4-androstene-3,17-dione (I.S.) was purchased from Zhuhai Yuancheng Pharmaceutical and Chemical Company. BSTFA and dimethyl aminopyridine (DMAP) were purchased from Sigma-Aldrich with purity >99% (confirmed by GC-MS).

MeCN and MeOH were of HPLC grade, while chloroform was of analytical grade and obtained from Fisher Scientific.

▶ Table 2 Retention times, LOD, and LOQ of the tested cannabinoids and the internal standard (IS).

Compound	Retention time	RTT	LOD (µg/mL)	LOQ (µg/mL)
CBDV	4.218	0.482	0.1	0.25
THCV	4.834	0.553	0.1	0.25
CBT	5.400	0.618	0.1	0.25
CBD	5.475	0.626	0.1	0.25
CBL	5.740	0.657	0.1	0.25
CBC	6.067	0.694	0.1	0.25
Δ ⁸ -THC	6.244	0.714	0.1	0.25
Δ ⁹ -THC	6.409	0.733	0.1	0.25
CBDVA	6.747	0.772	0.1	0.25
CBN	7.022	0.803	0.1	0.50
CBG	7.289	0.834	0.1	0.25
CBE	7.406	0.847	0.1	0.50
THCVA	7.876	0.901	0.1	0.25
CBDA	8.160	0.933	0.1	0.25
CBLA	8.384	0.959	0.1	0.50
THCAA	9.627	1.101	0.1	0.25
CBCA	9.839	1.125	0.1	0.25
CBGA	10.226	1.170	0.1	0.25
CBEA	10.406	1.190	0.1	0.50
CBNA	10.494	1.200	0.1	0.5
IS	8.743	1.000	-	-

RTT = Relative Retention Time; LOD = Limit of Detection; LOQ = Limit of Quantitation

Cannabis plant material

Cannabis plants (*in vitro* micropropagated at the University of Mississippi) were grown in an indoor controlled environment (light with a photoperiod of 16 h, 700 μ mol/m²/s, the temperature was set at 25–30 °C, and relative humidity (60%). The taxonomy of the plants was recognized by Dr. Suman Chandra and a voucher specimen (S1310V1) was kept at Coy–Waller Laboratory, School of Pharmacy, University of Mississippi, USA.

Instrumentation and column

GC-FID analysis was performed on Agilent 6890 Network GC System (Agilent Technologies) fitted with a 7683B-series injector. Separations were carried out on a DB-1MS column (15 m \times 0.25 mm, and 0.25 µm film thickness). Helium was used as the carrier gas at a flow rate of 0.8 mL/min and as detector make-up gas. The inlet temperature was set at 275 °C with a split ratio of 20:1. The temperature program started at 190 °C and was held for 1 min, then ramped to 230 °C with a rate of 30 °C/min (held for 2 min). Next, the oven temperature was ramped to 250 °C at a rate of 5 °C/min (held for 1 min), and then the temperature increased at a rate of 20 °C/min to 300 °C (held for 2.75 min). At the end of the gradient, the oven cooled back down

to 190 °C. The total run time was 17.5 min. Detector temperature was set at 300 °C and the hydrogen, air, and make-up flow rates were 40, 500, and 27 mL/min., respectively. Data were acquired and analyzed by Agilent ChemStation software (rev. B.04.02).

Internal standard and DMAP preparation

The IS (1 mg/mL of 4-androstene-3,17-dione) was prepared in MeOH and chloroform (9:1, v:v). A solution of 2% DMAP was prepared in MeOH.

Standard solutions preparation

From vials containing individual cannabinoids with a concentration of 1.0 mg/mL, 50 μ L were mixed. After evaporation to dryness under nitrogen, the cannabinoid mixture was re-dissolved in 1 mL of MeOH to get a stock standard solution of 50 μ g/mL. Serial dilutions were made to prepare the calibration curve points.

Calibration curves and control samples

A six-point calibration curve was prepared from stock standard solution (50 μ g/mL cannabinoid mixture) at 0.25–25 μ g/mL for all cannabinoids except CBN, CBL, CBE, and CBEA, which were prepared from the stock standard solution (50 μ g/mL cannabinoid

cman cman </th <th></th>																						
Under the field Desc Desc <thdesc< th=""> Desc Desc</thdesc<>		Concen-	Batch 1				Batch 2				Batch 3				Batch 4							
12.12.40100162.600.102.610.102.610.100.120.100.120.100.110.100.11<	0	tration (µg/mL)	Mean	SD	True- ness	Preci- sion (%RSD)	Mean	SD	True- ness	Preci- sion (%RSD)	Mean	SD	True- ness	Preci- sion (% RSD)	Mean	SD	True- ness	Preci- sion (%RSD)	Mean	SD	True- ness	Preci- sion (%RSD)
220001000		1.25	1.24	0.03	101%	2.60%	1.21	0.01	98%	1.00%	1.22	0.02	%66	1.20%	1.2	0.01	97%	1.10%	1.25	0.01	101%	1.10%
51011013203014015015017016 <td></td> <td>2.5</td> <td>2.5</td> <td>0.06</td> <td>101%</td> <td>2.50%</td> <td>2.45</td> <td>0.05</td> <td>%66</td> <td>2.20%</td> <td>2.4</td> <td>0.04</td> <td>97%</td> <td>1.70%</td> <td>2.49</td> <td>0.06</td> <td>100%</td> <td>2.40%</td> <td>2.5</td> <td>0.05</td> <td>101%</td> <td>2.20%</td>		2.5	2.5	0.06	101%	2.50%	2.45	0.05	%66	2.20%	2.4	0.04	97%	1.70%	2.49	0.06	100%	2.40%	2.5	0.05	101%	2.20%
11100		5	5.11	0.11	101%	2.20%	5.07	0.03	100%	0.60%	4.9	0.03	97%	0.70%	5.21	0.04	103%	0.80%	5.17	0.05	102%	0.90%
111		1.25	1.22	0.02	98%	1.20%	1.2	0.01	96%	1.20%	1.19	0.02	95%	1.30%	1.24	0.01	%66	1.10%	1.25	0.01	100%	1.10%
111		2.5	2.5	0.05	100%	2.10%	2.46	0.05	98%	2.10%	2.48	0.04	%66	1.80%	2.5	0.06	100%	2.30%	2.51	0.05	101%	2.10%
11100		ß	5.13	0.02	103%	0.50%	5.1	0.03	102%	0.60%	5.2	0.04	104%	0.80%	5.17	0.04	103%	0.80%	5.2	0.05	104%	0.90%
15163103103103104310		1.25	1.22	0.02	98%	1.30%	1.19	0.01	95%	1.10%	1.2	0.01	96%	1.20%	1.19	0.01	95%	1.20%	1.26	0.03	101%	2.40%
5 6 103 6305 503		2.5	2.5	0.05	100%	1.90%	2.45	0.05	98%	2.00%	2.48	0.04	%66	1.50%	2.48	0.05	%66	2.10%	2.5	0.05	100%	1.90%
1251210.029781.9081.900.910.871.9081.9100.910.931.0101.031.0100.930.0109.941252490.0519640.059872.1082.460.059872.1082.490.049872.400.059.950.050.051250.110.031030.500.510.031020.609872.1082.400.059.750.050.051250.120.031200.501200.501200.550.501200.550.560.550.560.750.560.750.750.760.750.760.751250.120.019881.1081.1081.1081.1081.1081.1081.1081.1081.1080.101250.120.130.109881.108<		ß	5.14	0.03	103%	0.50%	5.09	0.03	102%	0.60%	5.21	0.04	104%	0.80%	5.22	0.04	104%	0.80%	5.17	0.05	103%	%06.0
57249005100820082460059882108248046988240854005057050103103170110131038050851003103810381038103810381038103810381038175125010988120810381038103810381038103810381038103817512101098812081038103810381038103810381038103810381751210101038103811001010381100101038110811081108175121010103811081109110811091108110811081108110817512101010381108110910381109103811081108110811081751210101008110811091008110811081108110811081108175124010103812480101038124812481248124812481248124812481248125124010310481248124812481248124812481248124812481248124812481248124812481248124812481		1.25	1.21	0.02	97%	1.50%	1.19	0.01	95%	1.30%	1.19	0.02	95%	1.30%	1.16	0.01	93%	1.10%	1.23	0.01	%66	0.60%
		2.5	2.49	0.05	100%	2.00%	2.46	0.05	98%	2.10%	2.48	0.04	%66	1.80%	2.47	0.06	%66	2.40%	2.5	0.05	100%	2.10%
		ß	5.14	0.03	103%	0.50%	5.1	0.03	102%	0.60%	5.2	0.04	104%	0.70%	5.22	0.04	104%	0.80%	5.16	0.06	103%	1.20%
151510.051001002.460.06982.302.400.04982.500.040.05 <t< td=""><td></td><td>1.25</td><td>1.22</td><td>0.01</td><td>98%</td><td>1.20%</td><td>1.2</td><td>0.01</td><td>96%</td><td>1.10%</td><td>1.21</td><td>0.03</td><td>97%</td><td>2.70%</td><td>1.24</td><td>0.01</td><td>%66</td><td>1.00%</td><td>1.3</td><td>0.12</td><td>104%</td><td>9.10%</td></t<>		1.25	1.22	0.01	98%	1.20%	1.2	0.01	96%	1.10%	1.21	0.03	97%	2.70%	1.24	0.01	%66	1.00%	1.3	0.12	104%	9.10%
55160.02103%0.50%5.110.03102%0.50%5.220.04104%5.70%5.150.04103%5.190.09%5.190.05104%1.251.210.01100%1.20%1.10%1.		2.5	2.51	0.05	100%	2.00%	2.46	0.06	98%	2.30%	2.49	0.04	%66	1.70%	2.5	0.06	100%	2.40%	2.52	0.06	101%	2.40%
1.25 1.21 0.01 100° 1.20° 1.10°		ß	5.16	0.02	103%	0.50%	5.11	0.03	102%	0.60%	5.22	0.04	104%	0.70%	5.15	0.04	103%	0.90%	5.19	0.05	104%	0.90%
1.5 1.49 0.05 100 2.10 2.46 0.06 99 2.30 2.48 0.04 100 1.70 2.40 2.40 2.49 0.06 100 1.2 0.13 100 0.50 5.11 0.03 100 0.10 100 100 100 100 100 100 2.40 2.49 0.06 1.2 0.01 100 0.50 5.11 0.03 100 100 1.20 0.01 100 100 0.02 2.40 0.02 <		1.25	1.21	0.01	100%	1.20%	1.19	0.02	98%	1.30%	1.19	0.02	98%	1.50%	1.23	0.01	102%	1.10%	1.18	0.01	97 %	1.10%
510.03100%0.50%5.110.03100%5.110.035.150.04100%5.150.04100%5.290.06103%1.251.230.0398%1.2400.1299%1.40%1.230.0198%1.20%1.18%0.0294%2.550.06100%2.40%1.240.0299%1.40%1.23%0.0198%1.20%1.20%1.18%0.0294%2.550.06100%2.40%5.050.03101%0.60%5.130.04193%1.80%2.51%0.05100%2.49%0.06100%1.250.02103%0.40%5.050.03101%0.60%5.130.04103%0.70%5.150.05103%2.49%1.06%100%1.250.16103%0.10%1.20%0.10%1.20%0.10%1.20%1.20%100%100%1.251.240.02193%1.20%1.60%1.20%1.20%1.20%1.20%1.20%1.20%1.20%1.251.240.02193%1.20%1.20%1.20%1.20%1.20%1.20%1.20%100%1.20%100%1.251.240.02193%1.20%1.20%1.20%1.20%1.20%1.20%1.20%1.20%1.20%1.20%1.20%1.20%1.20%1.20%1.20%1.20%1.20%1.20%<		2.5	2.49	0.05	100%	2.10%	2.46	0.06	%66	2.30%	2.48	0.04	100%	1.70%	2.5	0.06	100%	2.40%	2.49	0.06	100%	2.30%
1.25 1.23 0.03 98% 1.24 0.01 99% 1.20% 1.21% 0.11 0.01 100% 1.20% 1.18 0.02 94% 2.5 0.06 100% 2.40% 2.47 0.06 99% 2.48 0.04 99% 1.80% 2.5 0.06 100% 2.49 0.06 100% 5 5.14 0.02 103% 0.03 101% 0.60% 5.13 0.04 193% 1.80% 2.5 0.06 100% 100% 100% 1.40% 1.05 100% 100% 2.40% 0.05 100% 103% 0.70% 5.15 0.05 100% 100% 100% 105% </td <td></td> <td>ß</td> <td>5.14</td> <td>0.03</td> <td>100%</td> <td>0.50%</td> <td>5.11</td> <td>0.03</td> <td>100%</td> <td>0.60%</td> <td>5.21</td> <td>0.04</td> <td>101%</td> <td>0.70%</td> <td>5.15</td> <td>0.04</td> <td>100%</td> <td>0.80%</td> <td>5.29</td> <td>0.06</td> <td>103%</td> <td>1.20%</td>		ß	5.14	0.03	100%	0.50%	5.11	0.03	100%	0.60%	5.21	0.04	101%	0.70%	5.15	0.04	100%	0.80%	5.29	0.06	103%	1.20%
2.5 2.6 100% 2.40% 0.47 0.06 99% 2.48 0.04 99% 1.80% 2.5 0.06 100% 2.40% 2.49 0.06 100% 5 5.14 0.02 103% 5.05 0.03 101% 0.60% 5.13 0.04 103% 0.70% 5.15 0.05 103% 5.26 0.05 105% 1.25 1.24 0.02 97% 1.30% 1.30% 1.26% 0.05 105% 105% 105% 105% 2.51 0.05 100% 1.22 0.02 97% 1.30% 1.25 0.05 105% 105% 105% 2.55 0.05 100% 2.54 0.05 1.26% 2.30% 2.30% 2.54 0.05 105% 105% 105% 105% 105% 105% 105% 105% 105% 105% 105% 105% 105% 105% 105% 105% 105% 105% 105%	U	1.25	1.23	0.03	98%	2.40%	1.24	0.02	%66	1.40%	1.23	0.01	98%	1.20%	1.24	0.01	100%	1.20%	1.18	0.02	94%	1.30%
5 5.14 0.02 103% 5.04 103% 5.13 0.04 103% 5.15 0.05 103% 5.26 0.05 105% 10		2.5	2.5	0.06	100%	2.40%	2.47	0.06	%66	2.30%	2.48	0.04	%66	1.80%	2.5	0.06	100%	2.40%	2.49	0.06	100%	2.50%
1.25 1.24 0.02 99% 1.50% 1.50% 1.30% 1.25 0.02 1.40% 1.26 0.02 101% 2.5 2.51 0.06 100% 2.48 0.06 99% 2.30% 2.55 0.05 100% 2.05 0.08 101% 3.10% 2.54 0.08 102% 5.15 0.03 103% 0.50% 5.14 0.03 103% 0.50% 5.24 0.04 105% 5.2 0.05 104% 5.22 0.05 104% 102% 5 5.14 0.03 103% 0.70% 5.24 0.04 105% 5.2 0.05 104% 5.22 0.05 104% 5.22 0.05 104% 0.52% 0.05 104% 0.52% 0.05 104% 0.52% 0.05 104% 0.52% 0.05 104% 0.52% 0.05 104% 0.52% 0.05 104% 0.05% 104% 0.05% 104% 0.05% <t< td=""><td></td><td>2</td><td>5.14</td><td>0.02</td><td>103%</td><td>0.40%</td><td>5.05</td><td>0.03</td><td>101%</td><td>0.60%</td><td>5.13</td><td>0.04</td><td>103%</td><td>0.70%</td><td>5.15</td><td>0.05</td><td>103%</td><td>0.90%</td><td>5.26</td><td>0.05</td><td>105%</td><td>0.90%</td></t<>		2	5.14	0.02	103%	0.40%	5.05	0.03	101%	0.60%	5.13	0.04	103%	0.70%	5.15	0.05	103%	0.90%	5.26	0.05	105%	0.90%
2.51 0.06 100% 2.40% 2.48 0.06 99% 2.30% 2.5 0.05 100% 2.00% 2.52 0.08 101% 3.10% 2.54 0.08 102% 5.15 0.03 103% 0.53 103% 0.70% 5.24 0.04 105% 0.80% 5.2 0.05 104% 0.90% 5.22 0.05 104% 0.05 104% 0.05 104% 0.05 104% 104% 0.05 104% 0.05 104% 104% 0.05 0.05 104% <t< td=""><td>υ</td><td>1.25</td><td>1.24</td><td>0.02</td><td>%66</td><td>1.40%</td><td>1.22</td><td>0.02</td><td>97%</td><td>1.60%</td><td>1.21</td><td>0.02</td><td>97%</td><td>1.30%</td><td>1.25</td><td>0.02</td><td>100%</td><td>1.40%</td><td>1.26</td><td>0.02</td><td>101%</td><td>1.40%</td></t<>	υ	1.25	1.24	0.02	%66	1.40%	1.22	0.02	97%	1.60%	1.21	0.02	97%	1.30%	1.25	0.02	100%	1.40%	1.26	0.02	101%	1.40%
5.15 0.03 103% 0.50% 5.14 0.03 103% 0.70% 5.24 0.04 105% 0.80% 5.2 0.05 104% 0.90% 5.22 0.05 104%		2.5	2.51	0.06	100%	2.40%	2.48	0.06	%66	2.30%	2.5	0.05	100%	2.00%	2.52	0.08	101%	3.10%	2.54	0.08	102%	3.00%
		J.	5.15	0.03	103%	0.50%	5.14	0.03	103%	0.70%	5.24	0.04	105%	0.80%	5.2	0.05	104%	%06.0	5.22	0.05	104%	1.00%

This document was downloaded for personal use only. Unauthorized distribution is strictly prohibited.

continued

Table 3 Intra-day precision and trueness (n = 30).

Com-	Concen-	Batch 1				Batch 2				Batch 3				Batch 4				Batch 5			
punod	tration (µg/mL)	Mean	SD	True- ness	Preci- sion (%RSD)	Mean	SD	True- ness	Preci- sion (%RSD)	Mean	SD	True- ness	Preci- sion (% RSD)	Mean	SD	True- ness	Preci- sion (%RSD)	Mean	OS -	True- ness	Preci- sion (%RSD)
CBDVA	1.25	1.15	0.02	92%	1.60%	1.2	0.02	86%	1.60%	1.13	0.02	91%	1.80%	1.14	0.02	91%	1.70%	1.2	0.02	896 %	1.50%
	2.5	2.41	0.07	97%	3.00%	2.41	0.07	86%	2.90%	2.39	0.05	896%	2.10%	2.38	0.07	95%	3.00%	2.45	0.07	98%	2.90%
	ß	5.09	0.03	102%	0.60%	4.98	0.03	100%	0.60%	5.18	0.04	104%	0.70%	5.2	0.04	104%	0.80%	5.1	0.05	102%	0.90%
CBN	1.25	1.21	0.01	97%	1.10%	1.19	0.02	95%	1.30%	1.24	0.01	%66	1.20%	1.25	0.01	100%	1.00%	1.25	0.01	100%	1.10%
	2.5	2.51	0.06	100%	2.30%	2.46	0.06	98%	2.30%	2.49	0.04	100%	1.80%	2.51	0.06	100%	2.50%	2.52	0.06	101%	2.40%
	Ŀ	5.15	0.03	103%	0.60%	5.1	0.03	102%	0.60%	5.16	0.04	103%	0.80%	5.15	0.05	103%	0.90%	5.19	0.05	104%	1.00%
CBG	1.25	1.23	0.06	98%	5.30%	1.19	0.02	95%	1.30%	1.2	0.02	96%	1.30%	1.23	0.01	98%	1.10%	1.21	0.01	97%	1.10%
	2.5	2.42	0.07	97%	3.00%	2.45	0.06	98%	2.50%	2.48	0.04	% 66	1.60%	2.52	0.06	101%	2.20%	2.52	0.07	101%	2.60%
	Ŀ	4.97	0.12	%66	2.50%	5.12	0.03	102%	0.60%	5.24	0.03	105%	0.60%	5.27	0.04	105%	0.80%	5.29	0.05	106%	0.90%
CBE	1.25	1.24	0.08	%66	6.50%	1.16	0.02	93%	1.30%	1.15	0.01	92 %	1.00%	1.16	0.01	93%	1.10%	1.17	0.01	93%	1.30%
	2.5	2.62	0.05	105%	1.80%	2.42	0.06	97%	2.40%	2.44	0.04	98%	1.80%	2.47	0.06	%66	2.60%	2.49	0.06	%66	2.50%
	ß	5.33	0.1	107%	1.90%	5.09	0.07	102%	1.30%	5.19	0.06	104%	1.20%	5.24	0.06	105%	1.20%	5.29	0.06	106%	1.20%
THCVA	1.25	1.17	0.02	94%	1.7%	1.15	0.01	92%	0.7%	1.17	0.02	93%	1.8%	1.22	0.01	98%	1.0%	1.21	0.02	97%	1.30%
	2.5	2.44	0.06	98%	2.3%	2.41	0.05	97%	2.2%	2.44	0.04	98%	1.7%	2.47	0.05	% 66	2.2%	2.46	0.05	98%	2.00%
	Ŀ	5.1	0.02	102%	0.4%	5.05	0.03	101%	0.6%	5.17	0.03	103%	0.6%	5.1	0.05	102%	0.9%	5.13	0.04	103%	0.90%
CBDA	1.25	1.19	0.02	95%	1.4%	1.16	0.02	93%	1.5%	1.22	0.01	97%	1.2%	1.22	0.01	98%	1.0%	1.23	0.01	98%	1.10%
	2.5	2.45	0.05	98%	2.1%	2.41	0.06	96%	2.3%	2.44	0.04	97%	1.7%	2.44	0.06	98%	2.3%	2.46	0.06	98%	2.30%
	2	5.1	0.02	102%	0.5%	5.04	0.03	101%	0.6%	5.07	0.03	101%	0.7%	5.05	0.04	101%	0.8%	5.09	0.05	102%	0.90%
CBLA	1.25	1.13	0.02	%06	1.6%	1.28	0.01	102%	1.0%	1.27	0.01	102%	0.9%	1.31	0.01	105%	1.1%	1.3	0.01	104%	1.1%
	2.5	2.45	0.06	98%	2.4%	2.51	0.05	100%	2.2%	2.51	0.04	101%	1.5%	2.55	0.06	102%	2.2%	2.54	0.06	102%	2.3%
	ß	5.13	0.02	103%	0.4%	5.05	0.03	101%	0.6%	5.14	0.04	103%	0.7%	5.14	0.04	103%	0.8%	5.16	0.05	103%	0.9%
THCAA	1.25	1.28	0.02	103%	1.50%	1.28	0.02	102%	1.20%	1.25	0.02	100%	1.80%	1.31	0.01	105%	0.70%	1.28	0.02	102%	1.50%
	2.5	2.52	0.06	101%	2.20%	2.49	0.06	100%	2.30%	2.48	0.04	%66	1.50%	2.51	0.07	100%	2.80%	2.51	0.06	100%	2.20%
	5	5.09	0.03	102%	0.60%	5.02	0.04	100%	0.80%	5.13	0.03	103%	0.70%	5.12	0.06	102%	1.20%	5.12	0.04	102%	0.80%

Gul W et al. Development and Validation ... Planta Med 2023; 89: 683-696 | © 2023. Thieme. All rights reserved.

continued

Out Bath I Bath I <th>Table</th> <th>Table 3 Continued</th> <th>7</th> <th></th>	Table	Table 3 Continued	7																			
Tation Frei <	Com-	Concen-	Batch 1				Batch 2				Batch 3				Batch 4				Batch 5			
1.25 1.19 0.02 100° 1.2 0.02 101° 1.30° 1.10° 1.30° 1.10° 1.30° 1.10° 1.30° 1.10° 1.30° 1.10° 1.30° 1.30° 1.10° 1.30° 1.30° 1.10° 1.01°	punod	tration (µg/mL)	Mean	S	True- ness	Preci- sion (%RSD)	Mean	SD	True- ness	Preci- sion (%RSD)	Mean	S	True- ness	Preci- sion (% RSD)	Mean	SD	True- ness	Preci- sion (%RSD)	Mean	SD	True- ness	Preci- sion (%RSD)
2.5 2.48 0.05 10.04 2.004 2.44 0.05 1.904 1.204 2.46 0.05 1.914 0.05 2.046 2.046 0.05 <t< td=""><td>CBCA</td><td>1.25</td><td>1.19</td><td>0.02</td><td>100%</td><td>1.60%</td><td>1.2</td><td>0.02</td><td>101%</td><td>1.30%</td><td>1.15</td><td>0.02</td><td>97%</td><td>1.50%</td><td>1.15</td><td>0.01</td><td>97%</td><td>1.30%</td><td>1.15</td><td>0.01</td><td>97%</td><td>1.10%</td></t<>	CBCA	1.25	1.19	0.02	100%	1.60%	1.2	0.02	101%	1.30%	1.15	0.02	97%	1.50%	1.15	0.01	97%	1.30%	1.15	0.01	97%	1.10%
5 5 6		2.5	2.48	0.05	100%	2.00%	2.44	0.05	%66	1.90%	2.45	0.03	%66	1.40%	2.45	0.05	%66	2.20%	2.46	0.05	100%	2.20%
1.25 1.17 0.02 9.48 1.406 1.15 0.01 224 1.06 1.16 0.10 1.208 <		ß	5.14	0.03	%66	0.60%	5.04	0.03	97%	0.60%	5.21	0.04	101%	0.80%	5.21	0.05	101%	%06.0	5.25	0.05	101%	1.00%
2.5 2.44 0.05 9.8° 2.10° 0.05 9.6° 2.43 0.05 2.45 0.05 9.8° 2.40° 2.49 0.05 2.49° 0.05° 2.49° 0.05° 2.49° 0.05°	CBGA	1.25	1.17	0.02	94%	1.40%	1.15	0.01	92 %	1.10%	1.15	0.02	92%	1.30%	1.14	0.01	91%	1.30%	1.25	0.02	100%	1.30%
5 6.08 0.02 102% 5.03 0.03 101% 0.50% 5.16 0.04 103% 5.18 0.05 104% 1.00% 5.12 0.06 102% 1.25 1.26 0.04 101% 3.00% 1.24 0.01 99% 1.20% 1.24 0.01 99% 1.20% 1.24 0.01 99% 2.55 0.06 102% 2.20% 2.49 0.05 98% 2.10% 2.49 0.06 102% 1.20% 10.0% 1.20% 1.20% 1.20% 10.0% 10.1% 10.1% 10.1% 10.1% 10.1% 10.1% 10.1% 10.1% 10.1% 10.1% 10.1% 10.1%<		2.5	2.44	0.05	98%	2.10%	2.4	0.05	86%	2.30%	2.43	0.04	97%	1.80%	2.45	0.06	98%	2.60%	2.49	0.06	100%	2.40%
1.25 1.26 0.04 101% 3.00% 1.24 0.01 99% 1.10% 1.23 0.01 99% 1.20% 1.24 0.01 99% 1.20% 1.24 0.01 99% 1.20% 1.20% 1.24 0.01 99% 1.20% 1.20% 1.24 0.01 99% 1.20% 1.20% 1.24% 0.01 109% 1.20%<		2	5.08	0.02	102%	0.40%	5.03	0.03	101%	0.50%	5.16	0.04	103%	%06.0	5.18	0.05	104%	1.00%	5.12	0.06	102%	1.20%
2:5 0:06 102% 2:40 0:05 98% 2:10% 0:44 0:45 0:46 0:06 100% 2:40% 2:57 0:06 100% 5 1:4 0:06 103% 1:20% 4:96 0:03 99% 0:50% 5.2 0:05 104% 1:0% 2:40% 5.21 0:07 104% 1:25 1:4 0:06 103% 1:20% 0:01 98% 1:10% 1:22 0:01 98% 1:01% 0:03 1:20% 5.21 0:01 98% 1:25 1:24 0:05 99% 1:10% 1:22 0:01 98% 1:10% 1:23 0:01 98% 1:04% 98% 1:10% 1:28 0:01 98% 1:04% 98% 1:10% 1:28 0:01 98% 1:01% 98% 1:03% 0:01 98% 0:01 98% 0:01 98% 0:01 98% 0:01 98% 0:01 98% 0:05	CBEA	1.25	1.26	0.04	101%	3.00%	1.24	0.01	%66	1.00%	1.22	0.01	98%	1.10%	1.23	0.01	%66	1.20%	1.24	0.01	%66	1.20%
5 5.14 0.06 103% 1.20% 4.96 0.03 99% 0.50% 5.2 0.05 104% 1.00% 5.16 0.06 103% 5.21 0.07 104% 1.25 1.24 0.02 99% 1.40% 1.22 0.01 98% 1.10% 1.23 0.01 98% 1.10% 123 0.01 98% 101% 98% 1.10% 123 0.01 98% 101% 98% 101% 98% 101% 123 0.01 98% 101% 98% 101% 103% 108% 101% 103% 0.01 98% 101% 98% 101% 103% 103% 101% 103% 101% 103%		2.5	2.55	0.06	102%	2.20%	2.44	0.05	98%	2.10%	2.48	0.04	%66	1.50%	2.49	0.06	100%	2.40%	2.5	0.06	100%	2.40%
1.25 1.24 0.02 99% 1.40% 1.22 0.01 98% 1.10% 1.23 0.01 98% 1.10% 1.23 0.01 98% 1.10% 1.23 0.01 98% 1.10% 1.23 0.01 98% 1.01% 98% 1.10% 1.23 0.01 98% 1.01 98% 0.03 0.03		ß	5.14	0.06	103%	1.20%	4.96	0.03	%66	0.50%	5.2	0.05	104%	1.00%	5.16	0.06	103%	1.20%	5.21	0.07	104%	1.30%
2.48 0.05 99% 2.20% 2.46 0.05 98% 2.10% 2.47 0.04 99% 1.60% 2.48 0.05 248 0.05 99% 101% 0.05 99% 101% 0.05 99% 101% 0.80% 5.15 0.04 103% 0.80% 5.14 0.04 103% 0.05 103% 103% 103% 103% 0.105 103% 103% 103% 101% 0.05 103%	CBNA	1.25	1.24	0.02	%66	1.40%	1.22	0.01	98%	1.10%	1.22	0.01	88%	1.10%	1.23	0.01	98%	1.10%	1.23	0.01	98%	1.10%
5.07 0.03 101% 0.50% 5.15 0.04 103% 0.80% 5.14 0.90% 5.15 0.05 103%		2.5	2.48	0.05	%66	2.20%	2.46	0.05	98%	2.10%	2.47	0.04	%66	1.60%	2.48	0.05	%66	2.20%	2.48	0.05	%66	2.20%
		5	5.07	0.03	101%	0.50%	5.05	0.03	101%	0.60%	5.15	0.04	103%	0.80%	5.14	0.04	103%	%06.0	5.15	0.05	103%	1.00%

Table 4 Inter-day precision and accuracy (n = 30).

Cannabinoid	Concentration	Between Batch	es		
	(µg/mL)	Mean	SD	Trueness	Precision (%RSD)
CBDV	1.25	1.22	0.02	99%	1.7%
	2.5	2.47	0.04	100%	1.8%
	5	5.09	0.12	101%	2.4%
THCV	1.25	1.22	0.03	98%	2.1%
	2.5	2.49	0.02	100%	0.8%
	5	5.16	0.04	103%	0.9%
СВТ	1.25	1.21	0.03	97%	2.4%
	2.5	2.48	0.02	99%	0.8%
	5	5.17	0.05	103%	1.0%
CBD	1.25	1.20	0.03	96%	2.2%
	2.5	2.48	0.02	99%	0.6%
	5	5.16	0.05	103%	0.9%
CBL	1.25	1.234	0.04	99%	3.2%
	2.5	2.496	0.02	100%	0.9%
	5	5.166	0.04	103%	0.8%
CBC	1.25	1.2	0.02	99.0%	1.7%
	2.5	2.484	0.02	99.8%	0.6%
	5	5.18	0.07	100.8%	1.4%
Δ ⁸ -THC	1.25	1.22	0.03	98%	2.1%
	2.5	2.49	0.01	100%	0.5%
	5	5.15	0.08	103%	1.5%
Δ ⁹ -THC	1.25	1.24	0.02	99%	1.7%
	2.5	2.51	0.02	100%	0.9%
	5	5.19	0.04	104%	0.8%
CBDVA	1.25	1.16	0.03	93%	2.9%
	2.5	2.41	0.03	96%	1.1%
	5	5.11	0.09	102%	1.7%
CBN	1.25	1.23	0.027	98%	2.2%
	2.5	2.50	0.024	100%	1.0%
	5	5.15	0.032	103%	0.6%
CBG	1.25	1.21	0.018	97%	1.5%
	2.5	2.48	0.044	99%	1.8%
	5	5.18	0.134	103%	2.6%
CBE	1.25	1.18	0.036	94%	3.1%
	2.5	2.49	0.079	100%	3.2%
	5	5.23	0.093	105%	1.8%
THCVA	1.25	1.18	0.03	95%	2.5%
	2.5	2.44	0.02	98%	0.9%
	5	5.11	0.04	102%	0.9%

continued

This document was downloaded for personal use only. Unauthorized distribution is strictly prohibited.

ysis.

oven at 40 °C for 24 h and then powdered. Triplicates of the ground samples (100 mg each) were weighed into a centrifuge tube and each extracted with 10 mL of a MeCN:MeOH mixture

mixture). Then, 50 µL of I.S and 10 µL of a 2% DMAP solution were

added to each concentration of the cannabinoid mixture. Under a

flow of nitrogen gas at 50 °C, the points were evaporated to dry-

ness in GC vials. Next, the residue obtained was derivatized by

adding 100 μL of BSTFA, the vials were capped and placed in an

oven set at 70°C for 30 min. Afterward, the vials were cooled to

room temperature, the contents transferred to inserts, and 2 µL

injections were made into the GC-FID. Calibration curves were ob-

tained in six replicates and constructed by plotting the concentra-

tion versus peak area ratio (peak area of analyte/peak area of I.S.).

Quality control samples independently prepared at three different

concentrations (low, medium, and high) for each cannabinoid

were prepared in the same way every day and on five consecutive days (one batch every day). All cannabinoid standards, stock solu-

tions, and QC samples were stored at - 20 °C till the time of anal-

Preparation of extracts of C. sativa plant materials

(8:2) by sonication for 30 min. The mixture was centrifuged for 5 min. at 1252 xg and transferred into pre-labeled extraction tubes. Aliquots of 10 μ L, 50 μ L, and 100 μ L were transferred into pre-labeled GC vials. To each vial, 50 μ L of 1 mg/mL I.S. solution and 10 μ L of 2% DMAP were added and the solvents evaporated to dryness using a gentle flow of nitrogen gas at 50 °C. The residue was then silylated by adding 100 μ L of BSTFA, vortexed, and the capped vials were kept in a 70 °C oven for 30 min. The vials were then brought to room temperature and the contents were transferred to 100 μ L GC-vial inserts and analyzed by the GC-FID. The injection volume was 2 μ L.

Method validation

The GC-FID method validation included linearity, selectivity, the limit of detection (LOD), the limit of quantification (LOQ), trueness, and precision and was performed according to the International Conference on Harmonization (ICH) Tripartite Guideline for Validation of Analytical Procedures [46]. Trueness was measured by the standard addition method. The intra-day and inter-day were assessed using a series of measurements. Six-point standard calibration curves were used to evaluate linearity. Calibration graphs were constructed by plotting the peak area ratio (y) of

Cannabinoid	Concentration	Between Batches			
	(µg/mL)	Mean	SD	Trueness	Precision (%RSD)
CBDA	1.25	1.20	0.03	96%	2.4%
	2.5	2.44	0.02	97%	0.8%
	5	5.07	0.03	101%	0.5%
CBLA	1.25	1.26	0.07	101%	5.8%
	2.5	2.51	0.04	101%	1.6%
	5	5.12	0.04	103%	0.8%
THCAA	1.25	1.28	0.02	102%	1.7%
	2.5	2.502	0.02	100%	0.7%
	5	5.096	0.05	102%	0.9%
CBCA	1.25	1.17	0.02	98%	2.1%
	2.5	2.46	0.02	99%	0.6%
	5	5.17	0.08	100%	1.6%
CBGA	1.25	1.17	0.04	94%	3.8%
	2.5	2.44	0.03	98%	1.3%
	5	5.11	0.06	102%	1.2%
CBEA	1.25	1.24	0.01	99%	1.2%
	2.5	2.49	0.04	100%	1.6%
	5	5.13	0.10	103%	2.0%
CBNA	1.25	1.23	0.01	98%	0.7%
	2.5	2.47	0.01	99%	0.4%
	5	5.11	0.05	102%	0.9%

Gul W et al. Development and Validation ... Planta Med 2023; 89: 683–696 | © 2023. Thieme. All rights reserved.

Variety	High THC Chemovar (R0466) 1606-THC	Intermediate Chemovar (R0466) 1326-IM	Intermediate Chemovar (R0466) B5-1597-IM	High CBD Chemovar (R0466) 1594-CBD
CBDV	ND	ND	ND	0.04
THCV	ND	ND	ND	ND
CBT	ND	ND	ND	ND
CBD	ND	0.61	1.04	2.13
CBL	ND	ND	ND	ND
CBC	ND	0.08	0.06	0.18
∆ ⁸ -THC	bloq	ND	0.05	ND
Δ ⁹ -THC	1.11	0.55	0.66	0.07
CBDVA	ND	0.03	0.05	0.19
CBG	bLOQ	0.03	0.06	0.04
CBN	bloq	ND	ND	ND
CBE	ND	ND	ND	ND
THCVA	0.08	0.05	0.04	ND
CBDA	ND	6.50	12.74	11.44
CBLA	ND	0.09	0.06	ND
THCAA	18.96	5.12	6.80	0.33
CBCA	0.22	0.35	0.55	0.52
CBGA	0.49	0.41	0.63	ND
CBEA	ND	ND	ND	ND
CBNA	0.12	0.15	0.18	ND

▶ Table 5 Calculated concentrations from different Cannabis sativa chemovars (% w/w).

bLOQ = below Limit of Quantitation. ND = not detected.

each analyte to that of IS versus the analyte concentration (x) by injecting triplicates of each concentration. Linear regression with a 1/x weighting factor described the regression relationship. Linearity was considered satisfactory if the correlation coefficient (R^2) of the calibration was higher than 0.99.

LOD and LOQ were determined as LOD = $3.3\sigma/S$ and LOQ = $10\sigma/S$, where σ = standard deviation of the response of each cannabinoid and S = slope of the calibration curve of each cannabinoid.

To verify method precision, the relative standard deviation (% RSD) of each batch (intra-day precision) was calculated for five consecutive days (n = 6) and between batches (inter-day precision) (n = 30). Trueness was calculated as % recovery and precision (stated as RSD%). The intra-day and inter-day precisions were required to be equal to or less than 15%, and the trueness to be within \pm 15% recovery.

Contributors' Statement

Data collection: I. Shahzadi, S. W. Gul, E. A. Ibrahim, S. Chandra, H. Lata; design of the study: W. Gul, M. A. ElSohly, M. M. Radwan, E. A. Ibrahim, I. Shahzadi; statistical analysis: I. Shahzadi, S. W. Gul, E. A. Ibrahim; analysis and interpretation of the data: E. A. Ibrahim, S.W. Gul, I. Shahzadi, S. Chandra, H. Lata, M.M. Radwan; drafting the manuscript: E.A. Ibrahim, S.W. Gul, M.M. Radwan, W. Gul; critical revision of the manuscript: M.A. ElSohly, W. Gul, M.M. Radwan.

Acknowledgements

This work was partially funded by the National Institute on Drug Abuse (NIDA) contract #N01DA-15-7793. Shahbaz W. Gul, Allison Mahurin, and Kareem ElSohly are acknowledged for technical help and for valuable suggestions during experimental and manuscript preparation.

Conflict of Interest

The authors declare that they have no conflict of interest.

References

 Pathak S, Deruiter J, Ramesh S, Govindarajulu M, Almaghrabi M, Nadar R, Moore T, Agrawal DC, Dhanasekaran M. The Legality of Use and Consumption of Cannabis (Marijuana) in the United States of America. Singapore: Springer Link; 2022: 113

This document was downloaded for personal use only. Unauthorized distribution is strictly prohibited.

- [2] Pollastro F, Minassi A, Fresu LG. Cannabis phenolics and their bioactivities. Curr Med Chem 2018; 25: 1160–1185
- [3] Radwan MM, Chandra S, Gul S, ElSohly MA. Cannabinoids, phenolics, terpenes and alkaloids of cannabis. Molecules 2021; 26: 2774
- [4] Mazurkiewicz-Bełdzińska M, Zawadzka M. Use of cannabidiol in the treatment of epilepsy. Neurol Neurochir Pol 2022; 56: 14–20
- Sharma A, Szaflarski J. Cannabidiol decreases seizure severity and brain temperature in treatment-resistant epilepsy. (P1–8.004). Neurology 2022; 98: 3625
- [6] Sarma ND, Waye A, ElSohly MA, Brown PN, Elzinga S, Johnson HE, Marles RJ, Melanson JE, Russo E, Deyton L. Cannabis inflorescence for medical purposes: USP considerations for quality attributes. J Nat Prod 2020; 83: 1334–1351
- Johnson MS, Wallace JG. Genomic and chemical diversity of commercially available high-CBD industrial hemp accessions. Front Genet 2021; 12: 1160
- [8] Crini G, Lichtfouse E, Chanet G, Morin-Crini N. Applications of hemp in textiles, paper industry, insulation and building materials, horticulture, animal nutrition, food and beverages, nutraceuticals, cosmetics and hygiene, medicine, agrochemistry, energy production and environment: A review. Environ Chem Lett 2020; 18: 1451–1476
- [9] Salami SA, Martinelli F, Giovino A, Bachari A, Arad N, Mantri N. It is our turn to get cannabis high: Put cannabinoids in food and health baskets. Molecules 2020; 25: 4036
- [10] Oh DA, Parikh N, Khurana V, Smith CC, Vetticaden S. Effect of food on the pharmacokinetics of dronabinol oral solution versus dronabinol capsules in healthy volunteers. Clin Pharmacol 2017; 9: 9
- [11] Nielsen S, Germanos R, Weier M, Pollard J, Degenhardt L, Hall W, Buckley N, Farrell M. The use of cannabis and cannabinoids in treating symptoms of multiple sclerosis: a systematic review of reviews. Curr Neurol Neurosci Rep 2018; 18: 1–12
- [12] Food and Drug Administration. Highlights of prescribing information: Epidiolex (cannabidiol) oral solution. Accessed August 22, 2022 at: https://www.epidiolex.com/sites/default/files/pdfs/EPIDIOLEX_Full_ Prescribing_Information_04_16_2020.pdf
- [13] Food and Drug Administration. FDA approves first drug comprised of an active ingredient derived from marijuana to treat rare, severe forms of epilepsy. Accessed August 22, 2022 at: https://www.fda.gov/newsevents/press-announcements/fda-approves-first-drug-comprisedactive-ingredient-derived-marijuana-treat-rare-severe-forms
- [14] European Medicines Agency. Epidyolex. Accessed August 22, 2022 at: https://www.ema.europa.eu/en/medicines/human/EPAR/epidyolex
- [15] Lah TT, Novak M, Pena Almidon MA, Marinelli O, Žvar Baškovič B, Majc B, Mlinar M, Bošnjak R, Breznik B, Zomer R. Cannabigerol is a potential therapeutic agent in a novel combined therapy for glioblastoma. Cells 2021; 10: 340
- [16] Vigli D, Cosentino L, Raggi C, Laviola G, Woolley-Roberts M, De Filippis B. Chronic treatment with the phytocannabinoid Cannabidivarin (CBDV) rescues behavioural alterations and brain atrophy in a mouse model of Rett syndrome. Neuropharmacology 2018; 140: 121–129
- [17] Abioye A, Ayodele O, Marinkovic A, Patidar R, Akinwekomi A, Sanyaolu A. Δ9-Tetrahydrocannabivarin (THCV): A commentary on potential therapeutic benefit for the management of obesity and diabetes. J Cannabis Res 2020; 2: 1–6
- [18] Khodadadi H, Salles ÉL, Shin E, Jarrahi A, Costigliola V, Kumar P, Yu JC, Morgan JC, Hess DC, Vaibhav K. A potential role for cannabichromene in modulating TRP channels during acute respiratory distress syndrome. J Cannabis Res 2021; 3: 1–7
- [19] Udoh M, Santiago M, Devenish S, McGregor IS, Connor M. Cannabichromene is a cannabinoid CB2 receptor agonist. Br J Pharmacol 2019; 176: 4537–4547

- [20] Mandrioli M, Tura M, Scotti S, Gallina Toschi T. Fast detection of 10 cannabinoids by RP-HPLC-UV method in Cannabis sativa L. Molecules 2019; 24: 2113
- [21] Patel B, Wene D, Fan ZT. Qualitative and quantitative measurement of cannabinoids in cannabis using modified HPLC/DAD method. J Pharm Biomed Anal 2017; 146: 15–23
- [22] Brighenti V, Pellati F, Steinbach M, Maran D, Benvenuti S. Development of a new extraction technique and HPLC method for the analysis of nonpsychoactive cannabinoids in fibre-type Cannabis sativa L.(hemp). J Pharm Biomed Anal 2017; 143: 228–236
- [23] Citti C, Pacchetti B, Vandelli MA, Forni F, Cannazza G. Analysis of cannabinoids in commercial hemp seed oil and decarboxylation kinetics studies of cannabidiolic acid (CBDA). J Pharm Biomed Anal 2018; 149: 532– 540
- [24] Križman M. A simplified approach for isocratic HPLC analysis of cannabinoids by fine tuning chromatographic selectivity. Eur Food Res Technol 2020; 246: 315–322
- [25] Berman P, Futoran K, Lewitus GM, Mukha D, Benami M, Shlomi T, Meiri D. A new ESI-LC/MS approach for comprehensive metabolic profiling of phytocannabinoids in Cannabis. Sci Rep 2018; 8: 1–15
- [26] Dei Cas M, Casagni E, Saccardo A, Arnoldi S, Young C, Scotti S, de Manicor EV, Gambaro V, Roda G. The Italian panorama of cannabis light preparation: Determination of cannabinoids by LC-UV. Forensic Sci Int 2020; 307: 110113
- [27] Gul W, Gul SW, Chandra S, Lata H, Ibrahim EA, ElSohly MA. Detection and quantification of cannabinoids in extracts of cannabis sativa roots using LC-MS/MS. Planta Med 2018; 84: 267–271
- [28] Happyana N, Agnolet S, Muntendam R, Van Dam A, Schneider B, Kayser O. Analysis of cannabinoids in laser-microdissected trichomes of medicinal Cannabis sativa using LCMS and cryogenic NMR. Phytochemistry 2013; 87: 51–59
- [29] Merone G, Tartaglia A, Rossi S, Santavenere F, Bassotti E, D'Ovidio C, Rosato E, de Grazia U, Locatelli M, Del Boccio P. Fast LC–MS/MS screening method for the evaluation of drugs, illicit drugs, and other compounds in biological matrices. Talanta Open 2022; 5: 100105
- [30] Merone GM, Tartaglia A, Rossi S, Santavenere F, Bassotti E, D'Ovidio C, Bonelli M, Rosato E, De Grazia U, Locatelli M. Fast quantitative LC-MS/ MS determination of illicit substances in solid and liquid unknown seized samples. Anal Chem 2021; 93: 16308–16313
- [31] Merone G, Tartaglia A, Rossi S, Santavenere F, Bassotti E, D'Ovidio C, Bonelli M, Rosato E, de Grazia U, Zanardo A. Fast liquid chromatography-tandem mass spectrometry method for the simultaneous determination of phytocannabinoids in oily based preparations. J Pharm Biomed Anal 2021; 203: 114174
- [32] Deidda R, Dispas A, De Bleye C, Hubert P, Ziemons É. Critical review on recent trends in cannabinoid determination on cannabis herbal samples: From chromatographic to vibrational spectroscopic techniques. Anal Chim Acta 2021; 1209: 339184
- [33] Amirav A, Neumark B, Margolin Eren KJ, Fialkov AB, Tal N. Cannabis and its cannabinoids analysis by gas chromatography–mass spectrometry with Cold EI. J Mass Spectrom 2021; 56: e4726
- [34] Bakro F, Jedryczka M, Wielgusz K, Sgorbini B, Inchingolo R, Cardenia V. Simultaneous determination of terpenes and cannabidiol in hemp (Cannabis sativa L.) by fast gas chromatography with flame ionization detection. J Sep Sci 2020; 43: 2817–2826
- [35] Baranauskaite J, Marksa M, Ivanauskas L, Vitkevicius K, Liaudanskas M, Skyrius V, Baranauskas A. Development of extraction technique and GC/FID method for the analysis of cannabinoids in Cannabis sativa L. spp. santicha (hemp). Phytochem Anal 2020; 31: 516–521
- [36] Cardenia V, Toschi TG, Scappini S, Rubino RC, Rodriguez-Estrada MT. Development and validation of a Fast gas chromatography/mass spectrometry method for the determination of cannabinoids in Cannabis sativa L. J Food Drug Anal 2018; 26: 1283–1292

- [37] Hazekamp A, Fischedick J. Cannabis-from cultivar to chemovar. Drug Test Anal 2012; 4: 660–667
- [38] Hazekamp A, Tejkalová K, Papadimitriou S. Cannabis: From cultivar to chemovar II–a metabolomics approach to Cannabis classification. Cannabis Cannabinoid Res 2016; 1: 202–215
- [39] Fodor B, Molnár-Perl I. The role of derivatization techniques in the analysis of plant cannabinoids by gas chromatography mass spectrometry. Trends Analyt Chem 2017; 95: 149–158
- [40] Fodor B, Boldizsár I, Molnár-Perl I. Alkylsilyl speciation and direct sample preparation of plant cannabinoids prior to their analysis by GC-MS. Anal Chim Acta 2018; 1021: 51–59
- [41] Leghissa A, Hildenbrand ZL, Foss FW, Schug KA. Determination of cannabinoids from a surrogate hops matrix using multiple reaction monitoring gas chromatography with triple quadrupole mass spectrometry. J Sep Sci 2018; 41: 459–468

- [42] Ye L, Budge SM. Sample preparation for the analysis of key metabolites from cannabinoids biosynthesis in phytoplankton using gas chromatography-mass spectrometry. J Am Oil Chem Soc 2022; 99: 115–127
- [43] Karschner EL, Barnes AJ, Lowe RH, Scheidweiler KB, Huestis MA. Validation of a two-dimensional gas chromatography mass spectrometry method for the simultaneous quantification of cannabidiol, Δ9-tetrahydrocannabinol (THC), 11-hydroxy-THC, and 11-nor-9-carboxy-THC in plasma. Anal Bioanal Chem 2010; 397: 603–611
- [44] Ahmed AQ, Noshad D, Li PC. Quantification of cannabinoids in cultivars of cannabis sp. by gas chromatography–mass spectrometry. Chromatographia 2021; 84: 711–717
- [45] Ibrahim EA, Gul W, Gul SW, Stamper BJ, Hadad GM, Salam RAA, Ibrahim AK, Ahmed SA, Chandra S, Lata H. Determination of acid and neutral cannabinoids in extracts of different strains of Cannabis sativa using GC-FID. Planta Med 2018; 84: 250–259
- [46] ICH Guidelines. Validation of analytical procedures: Text and methodology. Q2 (R1) 2005; 1: 05