

Accuracy of Flash Glucose Monitoring in Hemodialysis Patients With and Without Diabetes Mellitus



Authors

Michèle R. Weber¹, Matthias Diebold², Peter Wiesli¹, Andreas D. Kistler¹

Affiliations

- 1 Department of Medicine, Cantonal Hospital Frauenfeld, Frauenfeld, Switzerland
- 2 Clinic for Transplantation Immunology and Nephrology, University Hospital Basel, University of Basel, Basel, Switzerland

Key words

Hemodialysis, flash glucose monitoring, glucose monitoring, interstitial fluid glucose level, diabetes mellitus, FreeStyle Libre

received 16.02.2022

revised 27.04.2022

accepted 13.10.2022

published online 18.01.2023

Bibliography

Exp Clin Endocrinol Diabetes 2023; 131: 132–141

DOI 10.1055/a-1978-0226

ISSN 0947-7349

© 2023. The Author(s).

This is an open access article published by Thieme under the terms of the Creative Commons Attribution-NonDerivative-NonCommercial-License, permitting copying and reproduction so long as the original work is given appropriate credit. Contents may not be used for commercial purposes, or adapted, remixed, transformed or built upon. (<https://creativecommons.org/licenses/by-nc-nd/4.0/>)

Georg Thieme Verlag KG, Rüdigerstraße 14,
70469 Stuttgart, Germany

Corresponding

PD Dr. med. Andreas D. Kistler

Department of Medicine, Cantonal Hospital Frauenfeld

Pfaffenholzstrasse 4

8501 Frauenfeld

Switzerland

Tel.: +41 (0)52 723 76 43

andreas.kistler@stgag.ch

 **Supplementary Material** is available under <https://doi.org/10.1055/a-1978-0226>

ABSTRACT

Aims Glucose and insulin metabolism are altered in hemodialysis patients, and diabetes management is difficult in these patients. We aimed to validate flash glucose monitoring (FGM) in hemodialysis patients with and without diabetes mellitus as an attractive option for glucose monitoring not requiring regular self-punctures.

Methods We measured interstitial glucose using a FreeStyle Libre device in eight hemodialysis patients with and seven without diabetes mellitus over 14 days and compared the results to simultaneously performed self-monitoring of capillary blood glucose (SMBG).

Results In 720 paired measurements, mean flash glucose values were significantly lower than self-measured capillary values (6.17 ± 2.52 vs. 7.15 ± 2.41 mmol/L, $p = 1.3 \times 10^{-86}$). Overall, the mean absolute relative difference was 17.4%, and the mean absolute difference was 1.20 mmol/L. The systematic error was significantly larger in patients without vs. with diabetes (-1.17 vs. -0.82 mmol/L) and on dialysis vs. interdialytic days (-1.09 vs. -0.90 mmol/L). Compared to venous blood glucose (72 paired measurements), the systematic error of FGM was even larger (5.89 ± 2.44 mmol/L vs. 7.78 ± 7.25 mmol/L, $p = 3.74 \times 10^{-22}$). Several strategies to reduce the systematic error were evaluated, including the addition of $+1.0$ mmol/L as a correction term to all FGM values, which significantly improved accuracy.

Conclusions FGM systematically underestimates blood glucose in hemodialysis patients but, taking this systematic error into account, the system may be useful for glucose monitoring in hemodialysis patients with or without diabetes.

Introduction

Diabetes mellitus represents the leading cause of the end-stage renal disease (ESRD) [1] and accounts for about 20% of patients requiring renal replacement therapy in Europe [2]. In addition, due to other renal diseases, many patients with ESRD suffer from comorbid diabetes mellitus.

Adequate glycemic control in hemodialysis patients is hampered by a number of factors. First, advanced chronic kidney disease leads to altered insulin and glucose metabolism (reduced insulin secretion, insulin resistance, decreased insulin clearance, and reduced renal gluconeogenesis), which predisposes to both hyper- and hypoglycemia [3]. Second, the hemodialysis procedure itself can directly affect glucose and insulin levels [4]. Third, dialysis schedules often interfere with food intake; many dialysis patients suffer from gastrointestinal complaints, and they have to adhere to a variety of dietary restrictions. Fourth, most oral antidiabetic medications are contraindicated in ESRD. Finally, glycosylated hemoglobin (HbA1c) is not a reliable measure in dialysis patients due to reduced erythrocyte survival and treatment with erythropoiesis-stimulating agents (ESA). Given the unreliability of HbA1c and the frequent need for insulin treatment, many hemodialysis patients rely on self-monitoring of blood glucose (SMBG) using capillary blood. This is associated with further discomfort in patients who already experience regular punctures for the dialysis procedure.

In recent years, new devices have been introduced to measure interstitial glucose levels either continuously (continuous glucose monitoring, CGM) or on-demand (flash glucose monitoring, FGM) using a transcutaneous sensor [5–8]. In hemodialysis patients, interstitial glucose levels might correlate less well with blood glucose levels since volume overload is common and hemodialysis treatment induces rapid fluid shifts between different compartments. In addition, uremic toxins might interfere with the measurement method [9]. Several studies have been published recently that validated CGM in hemodialysis patients [4, 10–12]; in contrast, only very few data are available on FGM. Compared to CGM, FGM has several advantages: the sensor is cheaper and can be left in place for up to two weeks, as compared to 6 to 10 days for CGM [13, 14]. Furthermore, FGM devices are factory calibrated and do not require concomitant SMBG [15]. The most frequently used FGM system is the FreeStyle Libre (FGM, Abbott Diabetes Care, Alameda, CA) patch, which gets attached to the intact skin on the back of the upper arm. Interstitial glucose measurements are stored at 15-min intervals and can be displayed up to eight hours back after scanning the sensor with a reader [16].

Hemodialysis patients without diabetes also experience alterations in glucose- and insulin-metabolism and insulin sensitivity [3] with frequent hypoglycemic episodes [17]. Post-dialysis fatigue is a common complaint [18], and glucose-enriched dialysate has been found to decrease post-dialysis fatigue [19, 20], suggesting that fluctuating glucose levels in hemodialysis patients may contribute to symptoms. Thus, FGM might be also useful in monitoring the glucose profile of hemodialysis patients without diabetes, both on an individual basis and in clinical studies.

The purpose of this study was to evaluate the accuracy of the FreeStyle Libre FGM system to monitor glucose levels in both hemodialysis patients with and without diabetes mellitus.

Subjects, Materials, and Methods

Study Design

We performed a prospective observational study in two dialysis units run by the same hospital-based nephrology division (the Cantonal Hospital of Frauenfeld, Switzerland). The study was initiated in December 2019, and the last patient visit was in July 2020. The observation period for every patient started with the placement of a FreeStyle Libre FGM sensor and lasted for 14 days, corresponding to the lifetime of the sensor. The study was approved by the Ethics committee of Eastern Switzerland (Ethikkommission Ostschweiz, EKOS) and conducted in adherence to the Declaration of Helsinki. All patients gave written informed consent.

Study Population

All hemodialysis patients aged > 18 years of the two dialysis units who had been treated with chronic outpatient intermittent hemodialysis for at least three months were asked to participate in the study. Patients with impaired vision that might interfere with SMBG, reading FGM values, and completing a study diary were excluded. Further exclusion criteria were a history of allergic reaction to the material of FreeStyle Libre and regular intake of paracetamol as it potentially interferes with the measurement method [8, 9, 21]. The patient screening was based on patient history in the electronic health records.

Interventions and Measurements


A FreeStyle Libre FGM sensor was placed on the back of the upper arm of participants and left in place over the entire duration of the study. The attachment was always conducted by the same internal medicine resident (MRW), which had been instructed by a trained diabetic nurse prior to the investigation. The attachment always took place during the first dialysis session after the long interdialytic interval (i. e., on Monday or Tuesday). At the same time, patients were instructed for using the FreeStyle Libre FGM sensor and performing SMBG. Patients with diabetes continued to use their own SMBG devices, whereas patients without diabetes were instructed to use Contour XT. To ensure the highest amount of correspondence, the participants were asked to start the glucose measurements on the following day. Blood glucose levels were measured four times daily (before every meal and before bedtime) by SMBG, immediately followed by scanning of the FreeStyle Libre sensor. Participants were asked to record in a specially designed diary the readings as well as meals, physical activities, change of medication, or acute illness. The dialysate glucose concentration was 5.55 mmol/l. On the second and the third day of the trial (i. e., on a dialysis and a non-dialysis day), participants were instructed to perform three extra measurements within 1–2 h after every meal to measure postprandial glucose concentrations. In addition, we measured venous blood glucose values in the central hospital laboratory (on a Cobas pro integrated solution analyzer, Roche Diagnostics AG, Rotkreuz, Switzerland) immediately prior to every hemodialysis session.

After 14 days, participants returned the devices, and data from the FreeStyle Libre were downloaded using the FreeStyle Libre computer software. The readings in the diaries were verified by the authors, comparing them to the downloaded data. In the case of a

dislodged sensor, a new sensor was installed, and measurements were completed sticking to the priorly defined 14-day interval.

Statistical Analysis

We separately analyzed paired FGM and SMBG values obtained during the regular pre-prandial and nighttime measurements, the pre- and postprandial measurements from the second and third day of the study, and measurements obtained on dialysis and non-dialysis days. All participants with at least 50 % paired measurements available over a period of at least 7 days were included in the primary analysis. Participants with fewer paired measurements during a shorter period of time or with unreliable data due to non-compliance, acute illness, or prolonged use of paracetamol were excluded. We used Bland-Altman plots and paired t-tests to compare the difference between paired glucose values measured by SMBG vs. FGM. We further calculated the mean absolute difference (MAD) and the mean absolute relative difference (MARD) for each patient, for all measurements, and for subgroups of patients or measurement values, as defined. For the comparison of mean difference, MAD and MARD between dialysis vs. interdialytic days, between the first 5 days vs. the remainder of the study period, and between pre- vs. postprandial values, we used a t-test, and for the comparison between patients with vs. without diabetes and patients with high vs. low ultrafiltration volume, we used a nested t-test. Agreement between FGM and SMBG readings was further analyzed using the Surveillance error grid [22], available at <https://www.diabetestechology.org/seg/>. Statistical analyses were performed using Microsoft Excel version 2013 and R software, version 4.0.2 (R Core Team 2020. R: A language and environment for statistical computing. R Foundation for Statistical Computing, Vienna, Austria. URL <https://www.R-project.org/>).



Results

Patient Characteristics

Eighteen patients were included in the study. The median recording time of the sensor was 13 days (interquartile range 7–13). Three patients were excluded from the analysis: one discontinued due to acute illness requiring hospital admission, and two were non-compliant with SMBG (► **Fig. 1**). Data from 15 patients were analyzed, eight with and seven without diabetes mellitus. One FreeStyle Libre sensor had to be replaced during the study period due to sensor failure. The baseline characteristics of all patients included in the analysis are shown in ► **Table 1**, and characteristics, observation period, number of paired measurements, and measurement results for individual patients are shown in **Supplementary Table 1**.


Overall Accuracy of Flash Glucose Monitoring Results

In total, 795 paired glucose measurements were available for analysis, 720 of which were from the regular pre-prandial and bedtime

► **Table 1** Characteristics of study participants

Characteristics	All patients (n = 15)	Patients with diabetes (n = 8)	Patients without diabetes (n = 7)
Age (years)	71 ± 11	76 ± 7	65 ± 13
Sex (male)	11 (73)	5 (63)	6 (86)
Dry Weight (kg)	80.8 ± 15.8	87.9 ± 11.6	72.9 ± 16.8
BMI (kg/m ²)	28.6 ± 6.0	32.5 ± 4.0	24.5 ± 4.0
Dialysis vintage (months)	43 [22–48]	47 [38–67]	36 [14–45]
Urine volume (mL)	700 [200–1175]	920 [408–1163]	630 [75–1230]
Pre-dialysis creatinine level (µmol/L)	621 [514–702]	540 [500–662]	621 [584–768]
UF/W (mL/kg)	25 [4–32]	25 [15–28]	21 [4–32]
Vascular access (AV-fistula or graft/catheter)	12 (80) / 3 (20)	7(88) / 1 (13)	5 (71) / 2 (29)
Dialysis schedule, (morning/afternoon)	7 (47) / 8 (53)	5 (63) / 3 (38)	2 (29) / 5 (71)
Hemoglobin (g/L)	109 ± 12	104 ± 12	115 ± 9
Albumin (g/L)	39 ± 3	38 ± 2	39 ± 4
Duration of diabetes (years)		25.5 [20.5–27.0]	
HbA1c (%)		6.9 ± 0.5	
Treatment of diabetes (Insulin/OAD/none)		4 (50) / 3 (38) / 1 (13)	
Primary disease* (n)			
Diabetes mellitus	6 (40)	6 (75)	0 (0)
Hypertension	4 (26)	1 (13)	3 (43)
Obstructive uropathy	3 (20)	2 (25)	1 (14)
Glomerulonephritis	3 (20)	1 (13)	2 (29)
Unknown	2 (13)	1 (13)	1 (14)

Values are n (%), means ± SD or median [IQR]. *In some patients ESKD is multifactorial. Abbreviations: BMI, body mass index; UF/W, ultrafiltration volume per dialysis session (average of all dialysis sessions during the study period) divided by dry weight; OAD, oral antidiabetics.



► **Fig. 2** Glucose values for paired measurements by flash glucose monitoring (FGM) and self-monitoring blood glucose (SMBG) over the entire study period. Triangles and solid lines represent the mean of all FGM values, and crosses and dashed lines the mean of all SMBG results.

SMBG measurements and 75 from postprandial measurements. Mean values of all paired pre-prandial and nighttime glucose measurements by SMBG and FGM over the entire study period are shown in ► **Fig. 2**, mean glucose values over the entire study period by time of the day are shown in **Supplementary Figure 1**, and mean daily glucose profiles by diabetes status and hemodialysis vs. interdialytic days are mentioned in **Supplementary Figure 2**. Overall, mean FGM values were 0.98 mmol/L lower than SMBG values (6.17 ± 2.52 vs. 7.15 ± 2.41 mmol/L, $p = 1.3 \text{ E-}86$). Bland-Altman plots and Surveillance error grids for the entire study population, as well as by diabetes status, are shown in ► **Fig. 3**; 68.8% of FGM values fell into zone A, and 19.3% into zone B. MAD and MARD were 1.20 ± 0.92 mmol/L and $17.4 \pm 11.9\%$, respectively. The accuracy of the measurements varied considerably between patients (**Supplementary Figure 3**) but was relatively independent of blood glucose level (**Supplementary Figure 4** and **Supplementary Table 2**).


For the comparison of FGM values with venous blood glucose values, 72 pre-dialysis measurements were available. Mean FGM values were 1.87 mmol/L lower than venous BG values (5.89 ± 2.44 mmol/L vs. 7.78 ± 7.25 mmol/L, $p = 3.74\text{E-}22$), and 44.4% of FGM values were in zone A, whereas 51.4% fell into zone B of the

Surveillance error grid (**Supplementary Figure 5**). MAD and MARD were 2.0 ± 0.99 mmol/L and $26.3 \pm 12.39\%$, respectively.

Accuracy of Flash Glucose Monitoring Values by Patient Characteristics and Measurement Time

We next analyzed whether diabetes status and other variables had any influence on systematic and random measurement error. The systematic error was significantly larger in patients without vs. those with diabetes mellitus, during dialysis vs. interdialytic days, as well as with growing use time of the sensor, and tended to be higher in patients with vs. those without fluid removal during hemodialysis (► **Table 2**). However, these differences were rather small compared to the overall systematic measurement error.

Patients performed postprandial SMBG on both, dialysis and non-dialysis days. The results of paired pre- vs. postprandial measurements performed on these two days are shown in ► **Table 2** and the corresponding Bland Altman plots and Surveillance error grids are in **Supplementary Figure 6**.



► **Fig. 3** Bland-Altman Plots & Surveillance Error Grids for paired glucose measurements. Solid lines represent the mean difference and dashed lines the upper and lower confidence interval in the Bland-Altman Plots. Risk levels in the Surveillance Error Grids are color-coded and range from none (dark green) over moderate (yellow) to extreme (dark red).

Correction for the Systematic Bias of Flash Glucose Monitoring

Given that we found a systematic underestimation of blood glucose monitoring by FGM, we aimed to derive a correction term that might increase its accuracy in hemodialysis patients. Since the overall systematic bias was very close to 1 mmol/L, relatively independent of absolute glucose levels (**Supplemental Table 2**), and the effect of patient subgroup and measurement time was relatively small (► **Table 2**), we tested whether the simple correction term of +1 mmol/L to glucose readings obtained by FGM would result in a meaningful improvement of accuracy. After this correction, the

mean difference, MAD and MARD for FGM vs. SMBG in all 720 preprandial or bedtime paired measurements were 0.02 ± 1.15 mmol/L, 0.82 ± 0.81 mmol/L and $11.9 \pm 10.7\%$, respectively. Bland-Altman plots and Surveillance error grids for corrected FGM measurement values compared to SMBG are shown in ► **Fig. 4**; 87.8% of FGM values fell into zone A, and 10.4% into zone B. We also evaluated whether individual calibration of the device might improve accuracy. An individual correction term for every patient, based on the first four paired measurements, resulted in a mean difference, MAD and MARD of -0.51 ± 1.11 mmol/L, 0.89 ± 0.84 mmol/L, and $12.7 \pm 10.8\%$, respectively. An individual correction term based on the first two meas-

► **Table 2** Accuracy of flash glucose monitoring (FGM) by patient characteristics and measurement time.

	Paired Measurements (n)	FGM, mean ± SD (mmol/L)	SMBG, mean ± SD (mmol/L)	Mean difference (mmol/L)	MAD (mmol/L)	MARD (%)
Diabetes status						
Diabetes	392	7.5 ± 2.6	8.3 ± 2.5	-0.8 ± 1.3	1.1 ± 1.0	14.0 ± 10.7
No diabetes	328	4.5 ± 1.1	5.7 ± 1.2	-1.2 ± 1.0	1.3 ± 0.8	21.5 ± 11.9
<i>p-value</i>				0.00018	0.520	0.042
Fluid removal						
no / minimal	258	5.6 ± 2.0	6.4 ± 2.0	-0.8 ± 0.8	1.0 ± 0.6	15.7 ± 10.2
yes	462	6.5 ± 2.7	7.6 ± 2.5	-1.1 ± 1.3	1.3 ± 1.0	18.3 ± 12.6
<i>p-value</i>				0.366	0.212	0.507
Dialysis schedule						
Dialysis days	288	6.2 ± 2.7	7.3 ± 2.6	-1.1 ± 1.3	1.3 ± 1.0	18.6 ± 12.8
Interdialytic days	432	6.1 ± 2.4	7.0 ± 2.2	-0.9 ± 1.1	1.1 ± 0.8	16.6 ± 11.1
<i>p-value</i>				0.029	0.0047	0.002
Study period						
First 5 days	278	6.4 ± 2.7	7.1 ± 2.4	-0.8 ± 1.1	1.1 ± 0.8	15.6 ± 10.9
Day 6 – end	442	6.0 ± 2.4	7.2 ± 2.4	-1.1 ± 1.2	1.3 ± 1.0	18.5 ± 12.3
<i>p-value</i>				0.000037	0.0021	0.0018
Relation to meals						
Preprandial	86	5.6 ± 2.1	6.4 ± 1.8	-0.7 ± 1.1	1.0 ± 0.8	16.9 ± 13.6
Postprandial	75	7.1 ± 2.5	7.9 ± 1.9	-0.9 ± 1.3	1.3 ± 0.9	17.1 ± 11.1
<i>p-value</i>				0.441	0.033	0.926

MAD, mean absolute difference, MARD, mean absolute relative difference; SMBG, self-monitoring blood glucose; SD, standard deviation. Fluid removal refers to the average ultrafiltration volume per dialysis session over the entire study period. Patients with no/minimal fluid removal had a sufficient residual diuresis and only a minimal ultrafiltration volume (<250 mL and <5 mL / kg per dialysis session) to account for dialysate infused during flushing procedures. Patients with fluid removal depended on ultrafiltration (> 1500 mL and > 18 mL/kg per dialysis session) to treat hypervolemia.

urement days resulted in a mean difference, MAD and MARD of -0.33 ± 1.06 mmol/L, 0.78 ± 0.80 mmol/L, and 11.0 ± 10.1 %, respectively.

Discussion


The main findings of our study are that FGM using the FreeStyle Libre sensor led to a systematic underestimation of blood glucose levels in chronic hemodialysis patients, yet after adding a simple correction term, measured values were reasonably accurate.

CGM has gained growing popularity in the treatment of type 1 (T1DM) and recently also in type 2 diabetes mellitus (T2DM) patients. Current devices accomplish MARD results between 10 and 20% [15, 23–25]. ISO 1597:2013 defines that at least 95% of measurements have to lie within ± 15 % of the reference glucose concentration for blood glucose values ≥ 5.6 mmol/L or within ± 0.83 mmol/L for blood glucose values < 5.6 mmol/L [23]. Measurement of glucose in interstitial tissue by CGM results in a physiological time lag of approximately 10–15 min compared to blood glucose levels [26]. Deviation from blood glucose levels is greater during rapid glucose changes [15, 23, 24, 27] and hypoglycemia [24, 28, 29], but this bias is outweighed by several benefits of CGM: an enormous increase in the number of measurements and visualization of trends of daily glucose fluctuation that help to anticipate hypo- and hyperglycemia, with most of the devices including a warning system to prevent both hypo- and hyperglyce-

mia [13], thus extending time in euglycemia or in the target range. Compared to CGM, FGM has the advantage of factory calibration, thus obviating the need for SMBG to calibrate the device; the other advantages are a longer lifespan of the sensor [15], compact and lightweight design [24], and lower cost.

Tight blood glucose control is not considered essential in hemodialysis patients who mostly suffer from advanced macrovascular complications. However, even reasonable glycemic control is often difficult to achieve in these patients. Several studies have validated CGM in hemodialysis patients and found a reasonable agreement with SMBG [4, 10–12]. The use of FGM in hemodialysis patients, in contrast, has been reported only very recently. FGM would be particularly useful in hemodialysis patients due to its factory calibration eliminating the need for SMBG. A total of six studies have reported the use of FGM in hemodialysis patients so far [30–35], however, only three of these reported validations against SMBG [33–35]. Compared to these studies, we found a somewhat higher but similar precision of FGM, as expressed by MAD and MARD.

The mean difference reflects a systematic error, and MAD and MARD reflect a combination of systematic and random measurement errors. We found a significantly and relevantly lower mean glucose level by FGM as compared to SMBG, which pointed to a systematic measurement error. The underestimation of glucose levels was relatively constant across the absolute range of blood glucose levels (i. e., not proportional to the glucose level), which suggested that an additive term (rather than a correction factor) might



► **Fig. 4** Bland-Altman Plots & Surveillance Error Grids for paired glucose measurements, after adding a correction term of + 1 mmol/L to all flash glucose monitoring (FGM) values.

be used to correct the systematic measurement error. Since the mean difference between paired FGM and SMBG values was very close to 1 mmol/L, we evaluated the utility of simply correcting FGM values by adding 1 mmol/L. Indeed, this correction term considerably improved the performance of FGM and yielded a MARD and results on Surveillance's error grid which were equivalent to the respective measures reported for non-dialysis patients with T1DM and T2DM [6, 25, 28, 36] and in agreement with current recommendations and standards for CGM (ISO 1597:2013). While our study is the first to evaluate a correction term for the systematic measurement bias of FGM in hemodialysis patients, all three studies that previously validated FGM in hemodialysis patients found a near-identical systematic bias (Yajima et al.: -19.9 mg/dL = -1.10 mmol/L, Matoba et al.: -18.4 mg/dL = -1.02 mmol/L;

Genua et al. not reported in numbers but visible from the correlation graph) [33–35]. An individual correction term derived for every patient from a number of paired measurements might allow for a more precise, individualized correction since the mean difference, MAD, and MARD between FGM and SMBG varied considerably between patients (Supplement Table 1). However, two consecutive calibration days were needed to improve accuracy over the use of the simple universal correction term, and this would require complicated calculations to be performed by the patients, resulting in only minor improvements.

The reason for the systematic measurement error of FGM in hemodialysis patients remains unclear. It is likely that alterations in volume status might influence equilibration between blood and interstitial fluid. This should, however, theoretically influence ran-

dom variation rather than introduce a systematic bias. In addition, the dialysis procedure itself might lead to fluid and electrolyte shifts. We tested these hypotheses by separately analyzing days with vs. without dialysis treatment and patients with relevant fluid removal during hemodialysis vs. those with sufficient residual renal function not requiring relevant fluid removal. We found a somewhat higher systematic bias, MAD and MARD on dialysis days compared to non-dialysis days, while the difference between patients with relevant fluid removal vs. those without was not significant. In addition, measurement bias somewhat increased with the increasing use time of the sensors. However, these influences were not large compared to the overall systematic measurement error. Thus, other factors, such as interference by uremic toxins, likely contribute to the systematic underestimation of blood glucose levels by FGM. Further study will be required to evaluate whether the same systematic bias is also present in patients with advanced renal failure not treated by hemodialysis. The mentioned effects of ultrafiltration rate and dialysis day on the accuracy of FGM also need to be taken into account when interpreting individual patient data, but again, these influences on overall accuracy are relatively small.

We have used SMBG as the reference against which to compare FGM because the Freestyle Libre device is intended to be a replacement for SMBG, which represents the current standard of care. Furthermore, SMBG, as a comparator, provides more reference points and mirrors real-life accuracy during daily use, as compared to venous glucose measurements. In addition, we also compared FGM values to a smaller number of simultaneously performed venous blood glucose measurements and found an even higher underestimation of blood glucose by FGM compared to venous levels. Further study is needed to evaluate whether SMBG also leads to some systematic blood glucose underestimation in hemodialysis patients.

To our knowledge, our study is the first to evaluate FGM in hemodialysis patients without diabetes. While we found a somewhat reduced accuracy compared to patients with diabetes, overall accuracy was reasonable in patients without diabetes, particularly after considering the correction factor of +1 mmol/L. Hemodialysis patients without diabetes are at increased risk for hypoglycemic episodes due to a number of factors, e. g., reduced insulin clearance and reduced food intake [3, 37]. Thus, FGM might represent a valuable tool to evaluate hemodialysis patients without diabetes with suspected hypoglycemic episodes, and FGM might be used to evaluate the effect of subclinical hypoglycemic episodes on perceived dialysis fatigue [18], both on an individual patient level as well as in the setting of clinical studies. The higher systematic bias of FGM in hemodialysis patients without diabetes as compared to those with diabetes is not simply attributable to the lower accuracy of FGM at low glucose values since measurement error was largely independent of absolute blood glucose values. The use of different SMBG devices (Contour XT from Abbott by patients without diabetes vs. their own device by patients with diabetes) and patients without diabetes being less familiar with SMBG measurements are potential explanations. Probably more importantly, however, interindividual variability of measurement error was relatively high, and the number of patients in our study relatively low, such

that few patients without diabetes and with a high individual bias of their measurements may have influenced the overall comparison.

Our study has several limitations. First, the study population was relatively small. However, due to a comparatively long study period, it was possible to obtain an adequate amount of paired measurements. Second, patients with diabetes used their own SMBG devices due to optimal handling, and accuracy might depend on the device used. Third, since capillary glucose measurements were performed by the participants themselves, we cannot exclude technical errors by participants, although they were carefully instructed. However, the latter two limitations reflect the real-life setting of the study, which also represents a strength since FGM was validated under real-life conditions. Finally, relatively few paired venous blood glucose measurements were available as the gold standard against which to compare FGM.

Conclusion

Our study shows that FGM systematically underestimates blood glucose levels in hemodialysis patients. The systematic error depended on diabetes status and dialysis schedule, but these influences were relatively small compared to the overall systematic error of the entire cohort. Importantly, the accuracy of FGM can be improved in both patients with and without diabetes by correcting for systematic error, e. g., via a correction term of +1 mmol/L. Thus, our findings validate FGM as a potentially useful tool for glucose monitoring in hemodialysis patients, when considering and correcting for the systematic errors

Funding

This investigator-initiated study did not receive any external funding.

Author Contributions

MRW: investigation, formal analysis, visualization, writing – original draft; MD: formal analysis PW: conceptualization, resources, writing – review & editing, supervision; ADK: conceptualization, methodology, formal analysis, resources, data curation, writing – original draft, supervision, project administration

All authors have read and approved the final manuscript.

Acknowledgments

We thank the diabetes specialist nurses and the dialysis staff for their support of the study.

Conflict of interest

PW reports to have participated in the advisory board and received lecture fees from Abbott, Switzerland. The other authors have no conflict of interest to declare.

References

- [1] Koye DN, Magliano DJ, Nelson RG et al. The global epidemiology of diabetes and kidney disease. *Adv Chronic Kidney Dis* 2018; 25: 121–132. doi:10.1053/j.ackd.2017.10.011
- [2] ERA-EDTA Registry: ERA-EDTA Registry Annual Report 2019. Amsterdam UMC, location AMC, Department of Medical Informatics, Amsterdam, the Netherlands, 2021.
- [3] Rahhal M-N, Gharaibeh NE, Rahimi L et al. Disturbances in insulin – glucose metabolism in patients with advanced renal disease with and without diabetes. *J Clin Endocrinol Metab* 2019; 104: 4949–4966. doi:10.1210/jc.2019-00286
- [4] Sobngwi E, Ashuntantang G, Ndouma E et al. Continuous interstitial glucose monitoring in non-diabetic subjects with end-stage renal disease undergoing maintenance haemodialysis. *Diabetes Res Clin Pract* 2010; 90: 22–25. doi:10.1016/j.diabres.2010.06.001
- [5] Castellana M, Parisi C, Di Molfetta S et al. Efficacy and safety of flash glucose monitoring in patients with type 1 and type 2 diabetes: a systematic review and meta-analysis. *BMJ Open Diabetes Res Care* 2020; 8: e001092. doi:10.1136/bmjdr-2019-001092
- [6] Bailey T, Bode BW, Christiansen MP et al. The performance and usability of a factory-calibrated flash glucose monitoring system. *Diabetes Technol Ther* 2015; 17: 787–794. doi:10.1089/dia.2014.0378
- [7] Haak T, Hanaire H, Ajjan R et al. Flash glucose-sensing technology as a replacement for blood glucose monitoring for the management of insulin-treated type 2 diabetes: A multicenter, open-label randomized controlled trial. *Diabetes Ther* 2017; 8: 55–73. doi:10.1007/s13300-016-0223-6
- [8] Cappon G, Vettoretti M, Sparacino G et al. Continuous glucose monitoring sensors for diabetes management: A review of technologies and applications. *Diabetes Metab J* 2019; 43: 383–397. doi:10.4093/dmj.2019.0121
- [9] Hoss U, Budiman ES. Factory-calibrated continuous glucose sensors: The science behind the technology. *Diabetes Technol Ther* 2017; 19: 44–50. doi:10.1089/dia.2017.0025
- [10] Joubert M, Fourmy C, Henri P et al. Effectiveness of continuous glucose monitoring in dialysis patients with diabetes: The DIALYDIAB pilot study. *Diabetes Res Clin Pract* 2015; 107: 348–354. doi:10.1016/j.diabres.2015.01.026
- [11] Yeoh E, Lim BK, Fun S et al. Efficacy of self-monitoring of blood glucose versus retrospective continuous glucose monitoring in improving glycaemic control in diabetic kidney disease patients. *Nephrology* 2018; 23: 264–268. doi:10.1111/nep.12978
- [12] Riveline J-P, Teynie J, Belmouaz S et al. Glycaemic control in type 2 diabetic patients on chronic haemodialysis: Use of a continuous glucose monitoring system. *Nephrol Dial Transplant* 2009; 24: 2866–2871. doi:10.1093/ndt/gfp181
- [13] Klonoff DC, Ahn D, Drincic A. Continuous glucose monitoring: A review of the technology and clinical use. *Diabetes Res Clin Pract* 2017; 133: 178–192. doi:10.1016/j.diabres.2017.08.005
- [14] Buckingham B. Clinical overview of continuous glucose monitoring. *J Diabetes Sci Technol* 2008; 2: 300–306. doi:10.1177/193229680800200223
- [15] Mian Z, Hermayer KL, Jenkins A. Continuous glucose monitoring: Review of an innovation in diabetes management. *Am J Med Sci* 2019; 358: 332–339. doi:10.1016/j.amjms.2019.07.003
- [16] FreeStyle Libre. FreeStyle Libre System. n.d <https://www.freestyle.abbott/ch-de/produkte.html>
- [17] Gosmanov AR, Gosmanova EO, Kovcsdy CP. Evaluation and management of diabetic and non-diabetic hypoglycemia in end-stage renal disease. *Nephrol Dial Transplant* 2016; 31: 8–15. doi:10.1093/ndt/gfv258
- [18] Bossola M, Tazza L. Postdialysis fatigue: A frequent and debilitating symptom. *Semin Dial* 2016; 29: 222–227. doi:10.1111/sdi.12468
- [19] Raju SF, White AR, Barnes TT et al. Improvement in disequilibrium symptoms during dialysis with low glucose dialysate. *Clin Nephrol* 1982; 18: 126–129
- [20] Leski M, Niethammer T, Wyss T. Glucose-enriched dialysate and tolerance to maintenance hemodialysis. *Nephron* 1979; 24: 271–273. doi:10.1159/000181734
- [21] Feldman B, Brazg R, Schwartz S et al. Trial in patients with type 1 diabetes. *Diabetes Technol Ther* 2003; 5: 769–779
- [22] Klonoff DC, Lias C, Vigersky R et al. The surveillance error grid. *J Diabetes Sci Technol* 2014; 8: 658–672. doi:10.1177/1932296814539589
- [23] Freckmann G, Pleus S, Link M et al. In Accuracy of BG Meters and CGM Systems: Possible Influence Factors for the Glucose Prediction Based on Tissue Glucose Concentrations 2016; 31–42. doi:10.1007/978-3-319-25913-0_2
- [24] Fokkert MJ, van Dijk PR, Edens MA et al. Performance of the FreeStyle Libre Flash glucose monitoring system in patients with type 1 and 2 diabetes mellitus. *BMJ Open Diabetes Res Care* 2017; 5: e000320. doi:10.1136/bmjdr-2016-000320
- [25] Ólafsdóttir AF, Attvall S, Sandgren U et al. A clinical trial of the accuracy and treatment experience of the flash glucose monitor FreeStyle Libre in adults with type 1 diabetes. *Diabetes Technol Ther* 2017; 19: 164–172. doi:10.1089/dia.2016.0392
- [26] Freckmann G, Link M, Pleus S et al. Measurement Performance of Two Continuous Tissue Glucose Monitoring Systems Intended for Replacement of Blood Glucose Monitoring Parts of the data have previously been presented at the 77th Scientific Sessions of the American Diabetes Association in San Di. *Diabetes Technol Ther* 2018; 20: 541–549. doi:10.1089/dia.2018.0105
- [27] Chen C, Zhao X-L, Li Z-H et al. Current and emerging technology for continuous glucose monitoring. *Sensors* 2017; 17: 182. doi:10.3390/s17010182
- [28] Aberer F, Hajsek M, Rumpler M et al. Evaluation of subcutaneous glucose monitoring systems under routine environmental conditions in patients with type 1 diabetes. *Diabetes, Obes Metab* 2017; 19: 1051–1055. doi:10.1111/dom.12907
- [29] Galindo RJ, Migdal AL, Davis GM et al. Comparison of the FreeStyle Libre pro flash continuous glucose monitoring (CGM) system and point-of-care capillary glucose testing in hospitalized patients with type 2 diabetes treated with basal-bolus insulin regimen. *Diabetes Care* 2020; 43: 2730–2735. doi:10.2337/dc19-2073
- [30] Javherani R, Purandare V, Bhatt A et al. Flash glucose monitoring in subjects with diabetes on hemodialysis: A pilot study. *Indian J Endocrinol Metab* 2018; 22: 848. doi:10.4103/ijem.IJEM_520_18
- [31] Ushigome E, Matsusaki S, Watanabe N et al. Critical discrepancy in blood glucose control levels evaluated by glycated albumin and estimated hemoglobin A1c levels determined from a flash continuous glucose monitoring system in patients with type 2 diabetes on hemodialysis. *J Diabetes Investig* 2020; 11: 1570–1574. doi:10.1111/jdi.13286
- [32] Hu K, Peng H, Ma Y et al. Analysis of glycemic improvement in hemodialysis patients based on time in range, assessed by flash glucose monitoring. *Blood Purif* 2021; 50: 883–890. doi:10.1159/000513162
- [33] Matoba K, Hayashi A, Shimizu N et al. Comparison of accuracy between flash glucose monitoring and continuous glucose monitoring in patients with type 2 diabetes mellitus undergoing hemodialysis. *J Diabetes Complications* 2020; 34: 107680. doi:10.1016/j.jdiacomp.2020.107680

- [34] Yajima T, Takahashi H, Yasuda K. Comparison of interstitial fluid glucose levels obtained by continuous glucose monitoring and flash glucose monitoring in patients with type 2 diabetes mellitus undergoing hemodialysis. *J Diabetes Sci Technol* 2020; 14: 1088–1094. doi:10.1177/1932296819882690
- [35] Genua I, Sánchez-Hernandez J, Martínez M] et al. Accuracy of flash glucose monitoring in patients with diabetes mellitus on hemodialysis and its relationship with hydration status. *J Diabetes Sci Technol* 2021; 15: 1308–1312. doi:10.1177/1932296820975057
- [36] Bonora B, Maran A, Ciciliot S et al. Head-to-head comparison between flash and continuous glucose monitoring systems in outpatients with type 1 diabetes. *J Endocrinol Invest* 2016; 39: 1391–1399. doi:10.1007/s40618-016-0495-8
- [37] Gosmanov AR. A practical and evidence-based approach to management of inpatient diabetes in non-critically ill patients and special clinical populations. *J Clin Transl Endocrinol* 2016; 5: 1–6. doi:10.1016/j.jcte.2016.05.002