Initial trimming followed by complete removal of an esophageal self-expandable metal stent for stent-related symptoms







Fig. 2 The transected portion of the covered metal stent.



Fig. 3 The remainder of the covered metal stent following its complete removal 3 months later.

Video 2



Endoscopic trimming of the esophageal covered metal stent using argon plasma coagulation: the esophageal covered metal stent is seen protruding into the stomach wall; the distal part of the stent is trimmed using argon plasma coagulation in retroflexed view; the transected stent is removed using a snare.

Placement of long, protruding self-expandable metal stents (SEMSs) into the gastrointestinal lumen may cause related symptoms. A few reports have described the usefulness of argon plasma coagulation (APC) for trimming or fenestrating a



Endoscopic removal of the esophageal covered metal stent: the distal part of the stent is grabbed with a rat-toothed forceps, and the stent is removed completely using an inversion technique by rotating and withdrawing the endoscope.

SEMS [1–4]. We report a trimming technique for a covered SEMS in the esophagus using APC in a retrograde fashion, followed by its complete removal.

A 67-year-old woman presented with dysphagia. Esophagogastroduodenoscopy



Fig. 4 Esophagogram showing a patent esophagus with no extravasation of contrast.

(EGD) showed a large ulcerated tumor in the esophagus with tumor excavation. A 12-cm partially covered SEMS was placed across the tumor. Subsequently the patient was able to resume eating solid food and underwent chemotherapy. However, 1 month after stent placement, she developed epigastric pain and dysphagia from impaction of the stent into the proximal stomach (**>** Fig. 1 a). The distal portion of the stent was trimmed with APC using a generator at a setting of 80W and a flow rate of 2 L/min (**> Fig. 1 b; > Video 1**). The procedure was performed with the scope in a retroflexed position to prevent esophageal mucosal injury. A length of the stent (approximately 4cm) was completely severed in a circumferential manner and was successfully removed from the stomach (**S** Fig. 2). After the procedure, the patient's pain and dysphagia improved.

After 3 months, however, she developed severe acid reflux and we decided to remove the remainder of the stent. Hyperplastic tissue at the uncovered proximal part of the stent was leveled using a stiff snare and APC to free up some of the mesh from the mucosa. The distal part of the stent was then grabbed with a rattoothed forceps, and the endoscope was withdrawn in a steady rotational fashion, such that the mesh eventually inverted, was dislodged, and then was successfully removed en bloc (**> Fig.3; • Video 2**). A subsequent esophagogram demonstrated improvement of the stricture without evidence of contrast extravasation (**•** Fig.4). All of the patient's stent-related symptoms resolved after these interventions.

Endoscopy_UCTN_Code_TTT_1AO_2AZ

Competing interests: None

Takeshi Tsujino, John G. Lee, Kenneth J. Chang

Division of Gastroenterology and Hepatology, H.H. Chao Comprehensive Digestive Disease Center, University of California, Orange, California, USA

References

- 1 *Rerknimitr R, Naprasert P, Kongkam P* et al. Trimming a metallic biliary stent using an argon plasma coagulator. Cardiovasc Intervent Radiol 2007; 30: 534–536
- 2 *Ishii K, Itoi T, Sofuni A* et al. Endoscopic removal and trimming of distal self-expandable metallic biliary stents. World J Gastroenterol 2011; 17: 2652–2657
- 3 Hamada T, Nakai Y, Isayama H et al. Trimming a covered metal stent during hepaticogastrostomy by using argon plasma coagulation. Gastrointest Endosc 2013; 78: 817
- 4 *Tieu AH, Saxena P, Singh VK* et al. Fenestration of a covered metal stent during cystoduodenostomy using argon plasma coagulation. Endoscopy 2014; 46 (Suppl. 01): E512–E513

Bibliography

DOI http://dx.doi.org/ 10.1055/s-0042-102881 Endoscopy 2016; 48: E109–E110 © Georg Thieme Verlag KG Stuttgart · New York ISSN 0013-726X

Corresponding author *Kenneth J. Chang, MD*

H.H. Chao Comprehensive Digestive Disease Center, University of California Irvine Medical Center 101 The City Drive. Bldg. 22C Orange CA 92868 USA Fax: +1-714-456-7520 kchang@uci.edu