

JAAA CEU Program

Volume 28, Number 7 (July/August 2017)

Questions refer to Nair et al, “Verification of a Proposed Clinical Electroacoustic Test Protocol for Personal Digital Modulation Receivers Coupled to Cochlear Implant Sound Processors,” 625–635.

Learner Outcomes:

Readers of this article should be able to:

- Cite differences between Academy fitting guidelines for frequency modulated listening systems employed with hearing aids and Academy guidelines when fitting these systems to cochlear implant users.
 - Describe how electroacoustic measures can evaluate transparency of digital modulation listening systems combined with cochlear implant processors.
1. When fitting frequency modulation (FM) systems to hearing aids, Academy guidelines recommend use of:
 - a. behavioral testing
 - b. electroacoustic measures
 - c. both behavioral testing and electroacoustic measures
 2. Which frequencies are averaged to calculate transparency?
 - a. 500, 1000, and 2000 Hz
 - b. 750, 1000, and 2000 Hz
 - c. 750, 1000, and 1500 Hz
 3. The Schafer et al (2013) protocol utilizes a transparency measure of:
 - a. ± 1 dB
 - b. ± 2 dB
 - c. ± 3 dB
 4. Research that demonstrates digital modulation technology offers superior speech in noise recognition when compared with FM listening systems has been published by:
 - a. McShefferty et al (2015)
 - b. Schafer et al (2013)
 - c. Wolfe et al (2013)
 5. What is the manufacturer recommended EasyGain for Roger 14 and 17 receivers and the CI sound processors utilized in this study?
 - a. -10 dB
 - b. 0 dB
 - c. $+10$ dB
 6. The dedicated receiver used with the Cochlear N5/CP810 sound processor was the Phonak Roger:
 - a. 14
 - b. 16
 - c. 17
 7. The presentation level of the speech signal used in recording sound processor performance in Step 1 of the test protocol was:
 - a. 55 dB SPL
 - b. 60 dB SPL
 - c. 65 dB SPL
 8. Which CI sound processor tested in this study met transparency at default gain for both protocols 100 percent of the time?
 - a. Cochlear N5 (CP810)
 - b. Cochlear N6 (CP910)
 - c. Advanced Bionics Naida CI Q70
 9. For Advanced Bionics Naida CI Q70 sound processors coupled to Roger 17 DM receivers, the number of processors able to meet transparency after receiver gain was adjusted in both conditions was:
 - a. 11
 - b. 14
 - c. 16
 10. One limitation of this investigation cited by the research team was:
 - a. only two cochlear implant manufacturers' sound processors were evaluated
 - b. only universal receivers were evaluated
 - c. electroacoustic measurements were recorded on equipment unavailable to most clinicians



JAAA CEU PROGRAM

WHO? All members of the Academy receive the CE Registry as a member benefit and are eligible to participate in the JAAA CEU Program.

WHAT? The JAAA CEU Program offers a minimum of 1.6 CEUs (16 continuing education hours) per volume year. Individuals can submit one or all JAAA CEU assessments for scoring and CEU credit. Each JAAA assessment is worth .2 CEUs.

WHERE? *eAudiology.org—Your CEU Source*

Participants can complete the assessments using the eAudiology.org online submission system, which provides automatic feedback (score, correct answers) and automatic recording to the member's CE Registry record.

WHEN? Volume 28 (2017) assessments will be accepted through December 31, 2017. Volume 28 submissions will be accepted by e-mail online at eAudiology.org. Submissions are credited in the calendar year they are submitted.

Volume 27 (2016) assessments will be accepted for a separate registration fee of \$95 until December 31, 2017. You can earn up to 1.6 CEUs! To register visit eAudiology.org. Volume 27 (2016) assessments will only be accepted via the online program.

WHY? Because you want convenient and cost-effective CEUs!

HOW? To register online, go to www.eAudiology.org. Once you have registered, the JAAA CEU Program will be added to your dashboard, and you will be able to access the assessments from there. If submitting by mail, complete the following and send with your completed answer sheet to the address below.

Education Department, JAAA
American Academy of Audiology
11480 Commerce Park Drive, Suite 220
Reston, VA 20191

Name

Address

City State Zip Code

Telephone

E-mail Address

_____ Please enroll me for the volume 28 (2017) JAAA CEU Program. I am enclosing \$95 for the year.

_____ I am currently enrolled in the volume 28 (2017) JAAA CEU Program.

TOTAL AMOUNT ENCLOSED:

METHOD OF PAYMENT:

☐ Check # _____

Made payable to:
American Academy of Audiology, Inc.

Credit Card

- ☐ Visa
☐ MasterCard
☐ American Express
☐ Discover

Credit Card # _____

Exp. Date ____/____/____

TIER 1 CREDIT (For ABA certificants)

T1



Tier 1 credit is available in this issue of JAAA. In order to receive Tier 1 credit for this assessment, you must score 80% or better. The credits will appear on your Academy transcript as Tier 1.

☐ Please check here if you are seeking Tier 1 credit.