

Near-infrared spectroscopy

Virendra Jain, Hari H. Dash

Abstract

Tissue ischaemia can be a significant contributor to increased morbidity and mortality. Conventional oxygenation monitoring modalities measure systemic oxygenation, but regional tissue oxygenation is not monitored. Near-infrared spectroscopy (NIRS) is a non-invasive monitor for measuring regional oxygen saturation which provides real-time information. There has been increased interest in the clinical application of NIRS following numerous studies that show improved outcome in various clinical situations especially cardiac surgery. Its use has shown improved neurological outcome and decreased postoperative stay in cardiac surgery. Its usefulness has been investigated in various high risk surgeries such as carotid endarterectomy, thoracic surgeries, paediatric population and has shown promising results. There is however, limited data supporting its role in neurosurgical population. We strongly feel, it might play a key role in future. It has significant advantages over other neuromonitoring modalities, but more technological advances are needed before it can be used more widely into clinical practice.

Key words: Brain ischaemia, Near-infrared spectroscopy, Neuromonitoring

BACKGROUND

Measurement of tissue oxygenation is of paramount importance in critical care settings since tissue ischaemia can be a major contributor to morbidity and mortality. Estimation of global or systemic oxygenation with the help of parameters like pulse oximetry, blood gas analysis, mixed venous oxygen saturation are well established, but monitoring of regional tissue oxygen saturation is slowly gaining acceptance.^[1] Ump teen methods of measuring cerebral oxygenation are in vogue such as jugular venous oximetry, brain tissue oxygen tension, transcranial Doppler (TCD) and electroencephalogram (EEG). However, all these have significant limitations.^[2]

Department of Anaesthesiology and Pain Medicine, Fortis Memorial Research Institute, Gurgaon, Haryana, India

Address for correspondence:

Prof. Hari H. Dash, Department of Anaesthesiology and Pain Medicine, Fortis Memorial Research Institute, Gurgaon - 122 002, Haryana, India.
E-mail: dr.harihardash@gmail.com

Near-infrared spectroscopy (NIRS) is a non-invasive means of determining real-time changes in regional oxygen saturation (rSO_2) of cerebral and somatic tissues.^[3] Use of NIRS is slowly gaining popularity among physicians.

Jöbsis in 1977 initially observed that light in the NIR light spectrum (700–950 nm) can traverse biological tissue (myocardium and brain) because of the relative transparency of tissue to light in this wavelength range.^[4] NIRS devices rely on the Beer-Lambert law, which states that one can measure a concentration of a substance based on its absorption of light. The absorption is proportional to the concentration of certain chromophores, mainly iron in haemoglobin and copper in cytochrome aa3. In the brain, the primary light-absorbing molecules within the NIR range are metal complex chromophores: Oxyhaemoglobin, deoxyhaemoglobin and cytochrome c oxidase. Because the most haemoglobin in tissue is in the venous circulation, NIRS gives a venous-weighted relative oxygen index of tissue beneath the probe.^[5,6]

This is an open access article distributed under the terms of the Creative Commons Attribution-NonCommercial-ShareAlike 3.0 License, which allows others to remix, tweak, and build upon the work non-commercially, as long as the author is credited and the new creations are licensed under the identical terms.

For reprints contact: reprints@medknow.com

How to cite this article: Jain V, Dash HH. Near-infrared spectroscopy. J Neuroanaesthesiol Crit Care 2015;2:221-4.

Access this article online

Quick Response Code:



Website:

www.jnaccjournal.org

DOI:

10.4103/2348-0548.165045

PRINCIPLE

NIR light can be used to measure the regional cerebral tissue oxygen saturation. This technique uses principles of optical spectrophotometry. However, because of the poor signal-to-noise ratio as a result of the low intensity of transmitted light, most commercially available devices use reflectance-mode NIRS, in which receiving optodes are placed ipsilateral to the transmitter and exploit the fact that photons transmitted through a sphere will traverse an elliptical path in which the mean depth of penetration is proportional to the transmitter and receiver optode separation. Fundamental challenges posed in utilising transcranial reflectance NIRS to measure cerebral tissue oxygen saturation include the potential requirement for knowledge of the photon path length, the presence of non-haeme chromophores and variable light absorption by overlying extracerebral tissue.

MONITORS AVAILABLE

Worldwide many companies are marketing NIRS based devices to measure cerebral oximetry. But in India, two companies are marketing these devices. First is Sensmart X-100 by Nonin Medical Inc., (Minneapolis, MN, USA) and other is INVOS 3100 by Medtronic (Jacksonville, FL, USA) [Figure 1].

TECHNIQUE

NIRS sensors are applied on either side of forehead on a clean, dirt free, non-greasy skin. The sensors adhere to skin and evaluate the area underneath. The emitting and detecting devices are often referred to as the optodes. In adults, bilateral frontal cerebral oximetry is used to monitor perfusion to at risk areas of grey matter within cerebral cortex in the watershed areas between the anterior cerebral artery and middle cerebral artery. The smaller head circumference

of neonates and children permits greater depth of penetration and assessment of subcortical tissue oxygenation. The sensors illuminate up to a volume of 10 ml of hemispherical tissue. Light is generated at specific wavelengths typically by light-emitting diodes, and is usually detected by silicon photodiodes. Many systems incorporate 2 or more channels allowing monitoring of multiple tissue regions of interest simultaneously.

The radial depth will depend on the interoptode distance. The optodes are placed on one side of the forehead with an interoptode spacing of 4–7 cm. NIR time resolved spectroscopy with functional maximal optode spacing of 4 cm measure cerebral haemodynamic responses optimally and quantitatively [Figure 2]. The normal values of rSO₂ are reported to be 60–80%. A 20% decline in rSO₂ from baseline is considered to be an ischaemic threshold. In general, a decrease in rSO₂ is reflective of an increase in oxygen extraction and debt as a result of increased metabolism, decreased perfusion and/or stagnant perfusion. High rSO₂ may be indicative of increased perfusion, decreased tissue bed metabolism and/or less oxygen extraction.

CLINICAL APPLICATIONS

The proper management of brain oxygenation is one of the principal endpoints of all anaesthesia procedures, but the brain remains one of the least monitored organs during clinical anaesthesiology. There are some medical procedures where iatrogenic brain ischaemia is present, including carotid endarterectomy (CEA) in patients with high-grade carotid artery stenosis, temporary clipping in brain aneurysm surgery, hypothermic circulatory arrest for aortic arch procedures and others in which the pathology itself generates brain ischaemia, such as traumatic brain injury (TBI) and stroke.

Cardiac surgery

The most common application of NIRS is the assessment of cerebral oxygenation during and following



Figure 1: INVOS cerebral oximeter



Figure 2: Bilateral frontal optode placement

cardiac surgery. This is very useful in patients on cardiopulmonary bypass (CPB) where poor neurological outcome like stroke and postoperative cognitive dysfunction are a concern for all clinicians. NIRS use has been found to decrease cerebral desaturation events during CPB, fewer strokes, less postoperative major organ morbidity (mechanical ventilation, myocardial infarction).^[7,8] Similarly low rSO₂ levels below 50% have been associated with increased risk of postoperative cognitive dysfunction and prolonged hospital stay by about three-fold.^[9,10]

NIRS use is now widespread in cardiac surgery. In fact, NIRS use has been maximally adopted by the cardiac centres performing high-risk procedures such as aortic arch repair and deep hypothermic circulatory arrest. Many centres have now formulated algorithms for the clinical use of NIRS monitors in cardiac surgery patients.^[11]

Carotid endarterectomy

Various monitoring techniques like EEG, TCD, evoked potential monitoring have been used to predict cerebral ischaemia in CEA. NIRS being simple to use has also been extensively used to monitor cerebral perfusion. It can be a valuable tool to detect cerebral ischaemia during CEA. Various cut-off values like a decline of >20% from baseline was initially recommended^[12] but now various studies recommend a decrease in rSO₂ of more than 12% to be reliable, sensitive and reliable specific threshold for brain ischaemia.^[13]

Paediatrics

NIRS is being increasingly used both intraoperatively and postoperatively for paediatric cardiac surgery,^[14] neurosurgery and in the critical care setting for low birth weight infants, premature children at risk for apnoea.^[15] It also appears useful for measuring systemic perfusion via its somatic channels.

General anaesthesia

Although cardiovascular monitoring is extensively used in general anaesthesia, brain oxygenation monitoring use is generally limited. NIRS monitoring can help predict cerebral desaturation events in high-risk surgeries like shoulder surgery in beach chair position,^[16] thoracic surgery with one lung ventilation.^[17] Its use can be extended to optimise cerebral oxygenation in patients at risk of perioperative stroke.

Neurological patients

In patients with subarachnoid haemorrhage (SAH), episodes of angiographic vasospasms were associated with $24\% \pm 4\%$ rSO₂ reductions.^[18] In TBI, an association has been described between rSO₂ values <60% and mortality, intracranial hypertension and compromised

cerebral perfusion pressure (CPP).^[19] On comparison with SjvO₂ and PbtO₂ monitoring for measuring cerebral oxygenation, rSO₂ had low accuracy for detecting moderate cerebral hypoxia (PbtO₂ ≤ 15 mmHg) and was moderately accurate for detecting severe cerebral hypoxemia (PbtO₂ ≤ 12 mmHg).^[20] NIRS changes precede changes in ICP in patients having delayed traumatic haematomas.^[21]

Notably, cerebral oximetry is of limited value in TBI and SAH as cerebral oedema or haemorrhage largely affects NIRS signals, and hence should be used as an additional neuromonitoring tool. However, this disadvantage has been put to use for identifying intracranial haematomas^[22] and cerebral oedema.^[23] It has also been found useful to detect differences in cerebral haemodynamic responses of brain-injured patients to postural changes in the neurocritical care unit.^[24] NIRS has been used to test cerebral blood flow auto regulation.^[25] It has been employed to determine the lower threshold limit for auto regulation during both normothermia and hypothermia and thus, might help adjusting CPP in many cardiac arrest victims.^[26]

Advantages

- Non invasive
- Relatively easy to set up and use
- Lack of operator dependence
- High temporal resolution
- Can be used continuously in the intraoperative and postoperative period.

Disadvantages

- Regional monitoring technique
- Impossible to detect changes in areas located distant from the monitoring site
- Potential for contamination of signal by extracranial tissue and ambient light
- Limited spatial resolution
- Wide intra and interindividual baseline variability
- Best used as a trend monitor. Claims of the absolute threshold for cerebral ischaemia should be treated with caution.

LIMITATIONS AND CONFOUNDING FACTORS

- Presence of blood underneath the monitoring area
- Sampling volume comprises venous, capillary and arterial blood. For the cerebral cortex, average haemoglobin in tissues is distributed in proportion of 70 % venous and 30 % arterial. However, there can be significant biological variations in the blood making data interpretation difficult
- Non-haeme tissue chromophores such as melanin (hairs) or increased bilirubin in jaundice can confound rSO₂ values.

NIRS seems to be an important tool in clinician's armamentarium for predicting cerebral and somatic tissue desaturation events. Despite its potential advantages over other neuromonitoring techniques such as being user-friendly, non-invasive and measurements over multiple regions of interest simultaneously with high temporal resolution; further investigation and technological advances are necessary before it can be introduced more widely into clinical practice.

Financial support and sponsorship

Nil.

Conflicts of interest

There are no conflicts of interest.

REFERENCES

- Scheeren TW, Schober P, Schwarte LA. Monitoring tissue oxygenation by near infrared spectroscopy (NIRS): Background and current applications. *J Clin Monit Comput* 2012;26:279-87.
- Smith M. Perioperative uses of transcranial perfusion monitoring. *Neurosurg Clin N Am* 2008;19:489-502, vii.
- Highton D, Elwell C, Smith M. Noninvasive cerebral oximetry: Is there light at the end of the tunnel? *Curr Opin Anaesthesiol* 2010;23:576-81.
- Jöbsis FF. Noninvasive, infrared monitoring of cerebral and myocardial oxygen sufficiency and circulatory parameters. *Science* 1977;198:1264-7.
- Gupta N, Dash HH. Neuromonitoring. In: Mehta Y, editor. *Textbook of Critical Care*. 1st ed. New Delhi: Jaypee Publishers; 2015. p. 66-89.
- Murkin JM, Arango M. Near-infrared spectroscopy as an index of brain and tissue oxygenation. *Br J Anaesth* 2009;103 Suppl 1:i3-13.
- Zheng F, Sheinberg R, Yee MS, Ono M, Zheng Y, Hogue CW. Cerebral near-infrared spectroscopy monitoring and neurologic outcomes in adult cardiac surgery patients: A systematic review. *Anesth Analg* 2013;116:663-76.
- Murkin JM, Adams SJ, Novick RJ, Quantz M, Bainbridge D, Iglesias I, et al. Monitoring brain oxygen saturation during coronary bypass surgery: A randomized, prospective study. *Anesth Analg* 2007;104:51-8.
- Slater JP, Guarino T, Stack J, Vinod K, Bustami RT, Brown JM 3rd, et al. Cerebral oxygen desaturation predicts cognitive decline and longer hospital stay after cardiac surgery. *Ann Thorac Surg* 2009;87:36-44.
- de Tourney-Jetté E, Dupuis G, Bherer L, Deschamps A, Cartier R, Denault A. The relationship between cerebral oxygen saturation changes and postoperative cognitive dysfunction in elderly patients after coronary artery bypass graft surgery. *J Cardiothorac Vasc Anesth* 2011;25:95-104.
- Vohra HA, Modi A, Ohri SK. Does use of intra-operative cerebral regional oxygen saturation monitoring during cardiac surgery lead to improved clinical outcomes? *Interact Cardiovasc Thorac Surg* 2009;9:318-22.
- Samra SK, Dy EA, Welch K, Dorje P, Zelenock GB, Stanley JC. Evaluation of a cerebral oximeter as a monitor of cerebral ischemia during carotid endarterectomy. *Anesthesiology* 2000;93:964-70.
- Mille T, Tachimiri ME, Klerys C, Ticocelli G, Bellinzona G, Blangetti I, et al. Near infrared spectroscopy monitoring during carotid endarterectomy: Which threshold value is critical? *Eur J Vasc Endovasc Surg* 2004;27:646-50.
- Hayashida M, Kin N, Tomioka T, Orii R, Sekiyama H, Usui H, et al. Cerebral ischaemia during cardiac surgery in children detected by combined monitoring of BIS and near-infrared spectroscopy. *Br J Anaesth* 2004;92:662-9.
- Yamamoto A, Yokoyama N, Yonetani M, Uetani Y, Nakamura H, Nakao H. Evaluation of change of cerebral circulation by SpO₂ in preterm infants with apneic episodes using near infrared spectroscopy. *Pediatr Int* 2003;45:661-4.
- Murphy GS, Szokol JW, Marymont JH, Greenberg SB, Avram MJ, Vender JS, et al. Cerebral oxygen desaturation events assessed by near-infrared spectroscopy during shoulder arthroscopy in the beach chair and lateral decubitus positions. *Anesth Analg* 2010;111:496-505.
- Hemmerling TM, Bluteau MC, Kazan R, Bracco D. Significant decrease of cerebral oxygen saturation during single-lung ventilation measured using absolute oximetry. *Br J Anaesth* 2008;101:870-5.
- Bhatia R, Hampton T, Malde S, Kandala NB, Muammar M, Deasy N, et al. The application of near-infrared oximetry to cerebral monitoring during aneurysm embolization: A comparison with intraprocedural angiography. *J Neurosurg Anesthesiol* 2007;19:97-104.
- Dunham CM, Ransom KJ, Flowers LL, Siegal JD, Kohli CM. Cerebral hypoxia in severely brain-injured patients is associated with admission Glasgow Coma Scale score, computed tomographic severity, cerebral perfusion pressure, and survival. *J Trauma* 2004;56:482-9.
- Leal-Noval SR, Cayuela A, Arellano-Orden V, Marín-Caballos A, Padilla V, Ferrández-Millón C, et al. Invasive and noninvasive assessment of cerebral oxygenation in patients with severe traumatic brain injury. *Intensive Care Med* 2010;36:1309-17.
- Gopinath SP, Robertson CS, Contant CF, Narayan RK, Grossman RG, Chance B. Early detection of delayed traumatic intracranial hematomas using near-infrared spectroscopy. *J Neurosurg* 1995;83:438-44.
- Robertson CS, Gopinath SP, Chance B. A new application for near-infrared spectroscopy: Detection of delayed intracranial hematomas after head injury. *J Neurotrauma* 1995;12:591-600.
- Gill AS, Rajneesh KF, Owen CM, Yeh J, Hsu M, Binder DK. Early optical detection of cerebral edema *in vivo*. *J Neurosurg* 2011;114:470-7.
- Kim MN, Edlow BL, Durduran T, Frangos S, Mesquita RC, Levine JM, et al. Continuous optical monitoring of cerebral hemodynamics during head-of-bed manipulation in brain-injured adults. *Neurocrit Care* 2014;20:443-53.
- Steiner LA, Pfister D, Strebel SP, Radolovich D, Smielewski P, Czosnyka M. Near-infrared spectroscopy can monitor dynamic cerebral autoregulation in adults. *Neurocrit Care* 2009;10:122-8.
- Lee JK, Brady KM, Mytar JO, Kibler KK, Carter EL, Hirsch KG, et al. Cerebral blood flow and cerebrovascular autoregulation in a swine model of pediatric cardiac arrest and hypothermia. *Crit Care Med* 2011;39:2337-45.