Technical Note

Initial Clinical Experience of Using a Newly Designed Preshaped Microguidewire in Acute Endovascular Thrombectomy

Abstract

Mechanical thrombectomy has been widely used for the treatment of acute ischemic stroke. During this procedure, operators must navigate the microcatheter with a microguidewire (MGW) into vessels that cannot be visualized on fluoroscopy as rapidly as possible. In this study, we developed a modified pigtail-shaped MGW (MPMGW) for security and controllability. Moreover, the efficacy of the MPMGW for the treatment of acute ischemic stroke was assessed. The MPMGW was designed using 0.014 MGW. Because we created four MPMGWs during a clinical evaluation before the launch in the market, these wires were used in four consecutive patients with acute ischemic stroke in the single institution. The occluded arteries were the basilar artery (n = 1), middle cerebral arteries (M1 and M2, n = 2), and internal carotid artery (n = 1). All four procedures were conducted without any complications. The procedures included navigating the MGW and passing it through the clot. Complete recanalization was achieved in all cases. The average time between femoral artery puncture and recanalization was 15 min. The use of the preshaped MPMGW in acute thrombectomy was effective in terms of both security of procedure and reduction in recanalization time.

Keywords: Acute ischemic stroke, endovascular treatment, microguidewire, modified pigtail-shaped, preshape, thrombectomy

Tomotaka Ohshima¹, Reo Kawaguchi², Ryuya Maejima², Naoto Yukiue², Naoki Matsuo², Shigeru Miyachi^{1,2}

¹Neuroendovascular Therapy Center, Aichi Medical University, ²Department of Neurosurgery, Aichi Medical University, Nagakute, Aichi 480-1195, Japan

Introduction

Rapid recanalization is the most effective treatment for acute ischemic stroke due to major cerebral artery occlusion.^[1] Recently, endovascular technologies and techniques for acute ischemic stroke have continuously evolved. However, endovascular stent clot retrieval is still challenging to perform and has risks, particularly for beginners, because operators must navigate the microcatheter with a microguidewire (MGW) into far distal vessels that cannot be visualized on fluoroscopy as rapidly as possible. Some authors have reported the use of a modified pigtail-shaped MGW (MPMGW) for various endovascular treatments.[2-6] Although the MPMGW is effective for acute thrombectomy, the manual quick shaping of this wire requires some skills. The common measurement of the round tip is >3.0 mm. Therefore, we designed a preshaped MPMGW with a round tip measuring 2.5 mm. Herein, we report the initial clinical experience using preshaped MPMGW for the treatment of acute ischemic stroke.

This is an open access journal, and articles are distributed under the terms of the Creative Commons Attribution-NonCommercial-ShareAlike 4.0 License, which allows others to remix, tweak, and build upon the work non-commercially, as long as appropriate credit is given and the new creations are licensed under the identical terms.

Materials and Methods

Microguidewire

A preshaped MPMGW was designed using a 0.014 MGW (Eiger 0.014; Medical Innovation Co., Ltd., Tokyo, Japan) [Figure 1]. The tip was bent round, with a diameter of 2.5 mm, and a portion with a length of 2.0 mm was kept straight. At the 10.0 mm proximal side of the tip, the wire was bent at 135°. The round tip was designed to prevent vessel injury. The 2.0-mm straight portion facilitated the easy insertion of the wire into the microcatheter. The bent tip allowed operators to select branches with torque. The MGW was packed with an introducer sheath beforehand, as navigation into the microcatheter was quickly performed. Because we created four preshaped MPMGWs during a clinical evaluation before launching in the market, these wires were used in four consecutive patients with acute ischemic stroke in a single institution.

How to cite this article: Ohshima T, Kawaguchi R, Maejima R, Yukiue N, Matsuo N, Miyachi S. Initial clinical experience of using a newly designed preshaped microguidewire in acute endovascular thrombectomy. Asian J Neurosurg 2020;15:241-4. Submission: 17-12-2019 Accepted: 24-01-2020

Published: 25-02-2020

Address for correspondence: Dr. Tomotaka Ohshima, Neuroendovascular Therapy Center, Aichi Medical University, 1-1 Yazakokarimata, Nagakute, Aichi 480-1195, Japan. E-mail: tmtkoh@gmail.com



For reprints contact: reprints@medknow.com



Figure 1: Image of the preshaped modified pigtail-shaped microguidewire

Endovascular procedure

The ASAP technique^[7] is commonly used for the anterior circulation, and the simple stent retrieving is used for the posterior circulation. The guiding catheter, aspirator, microcatheter, and stent retriever were selected based on the operator's preference. The preshaped MPMGW was used in all procedures.

Results

A summary of the clinical data, devices used, and clinical outcomes are shown in Tables 1-3, respectively. The occluded arteries were the basilar artery (n = 1), middle cerebral arteries (M1 and M2, n = 2), and internal carotid artery (n = 1). The guiding catheters used were 6-Fr Envoy (90 cm; Codman, Raynham, MA, USA) for posterior circulation and 9-Fr Optimo (Tokai Medical Products, Aichi, Japan) for anterior circulation. Penumbra ACE 68 (Penumbra, Alameda, CA, USA) was used as an aspirator. Moreover, Rebar 18 and Marksman microcatheters (Medtronic, Minneapolis, MN, USA) were used. Revive (Codman) and Solitaire (Medtronic) stent devices were used as clot retrievers. The operative images of representative patients are shown in Figures 2 and 3.

All four procedures have been conducted without any complications, and the procedures involved navigating the MPMGW and passing it through the clot. Complete recanalization was achieved in all cases. The average time between femoral artery puncture and recanalization was 15 min.

Discussion

Some authors have previously reported that the use of an MPMGW is an extremely safe and versatile method for neuroendovascular procedures.^[2-6] The round tip of the wire prevents unintentional entry into the perforating branches or blebs of an aneurysm and reduces the risks of perforation if it enters them erroneously. Moreover, MPMGW is useful

Table 1: Summary of clinical data								
	Age/ Sex	Location	02A (min)	NIHSS	ASPTECTS	Af		
Case 1	87/M	BA	120	34	6	+		
Case 2	75/M	M2	100	20	9	+		
Case 3	42/F	M1	30	16	7	-		
Case 4	70/M	ICA	330	28	5	+		

BA – Basilar artery; M – Middle cerebral artery; ICA – Internal carotid artery; O2A – Onset to arrival time; NIHSS – National Institutes of Health Stroke Scale; CT – Computed tomography; ASPECTS – Alberta Stroke Program Early CT score

Table 2: Summary of applicated devices							
	GC	Aspirator	MC	MGW	stent		
Case 1	6F Envoy	-	Reabar 18	Eiger 14	Revive 4.5*22		
Case 2	9F Optimo	ACE 68	Rebar 18	Eiger 14	Solitaire 4*40		
Case 3	9F Optimo	ACE 68	Marksman	Eiger 14	Solitaire 4*40		
Case 4	9F Optimo	ACE 68	Marksman	Eiger 14	Solitaire 6*40		
GC – Guiding catheter; MC – Microcatheter;							
MGW – Microguidewire							

Table 3: Summary of clinical outcomes							
	tPA	Number of pass	TICI	P2R (min)	mRS 30		
Case 1	-	1	3	14	5		
Case 2	+	2	3	22	1		
Case 3	-	1	3	10	0		
Case 4	+	1	3	15	3		

TICI – Thrombolysis in cerebral infarction scale; P2R – Puncture to recanalization time

in acute thrombectomy. When a stent retriever is used, the MGW must pass through the occluded vessel, which is not observed during angiography. The MPMGW can be safely advanced across a lesion even if an unseen aneurysm is in the distal vessel. The wire also provides better vessel selectivity with torque. When a guide catheter cannot be advanced up to the common carotid artery, the MPMGW allows selection into the internal carotid artery and navigation into intracranial small arteries with the same tip shape. Moreover, this type of MGW facilitates better microcatheter trackability, because it acts as an anchor in the distal vessels compared to a conventional right-angled shape MGW.

Before we developed the preshaped MPMGW, we manually used Chikai 14 MGW (Asahi Intecc Co., Ltd., Aichi, Japan). Beginners often take a few minutes to shape the wire. Furthermore, making the small round tip is quite challenging for beginners and even experts. Therefore, the preshaped MPMGW was designed to facilitate a rapid procedure and recanalization.

In this case series, the MGW can be smoothly inserted into the microcatheter and navigated up to around the M2–M3 junction. The small round tip can always be navigated at the occluded area without any manipulations



Figure 2: Images of case 1. R indicates the right side. (a) Right vertebral angiographic image showing basilar trunk occlusion (arrow). (b) Tip of the preshaped microguidewire (arrow). (c) The stent retriever is deployed (arrowheads). (d) Complete recanalization has been achieved

and negotiations. Moreover, it never migrated into the branching and perforating arteries.

In cases no. 1 and no. 4, the tip of the MGW appeared bent during penetration of the clots, which were found relative hard. As previously reported,^[3] the configuration of the MGW provides information about the quality of the clot during penetration. The preshaped MPMGW was harder, and it can penetrate hard clots more smoothly than the Chikai 14 MGW without dislocation of clots distally. Since both patients presented with hard clots, the corkscrew penetrating method was used.^[4]

The present study had some limitations. That is, a small sample was included, the same operator conducted all of the procedures, and the study was performed in a single institution. Moreover, the time between puncture and recanalization in our study was shorter than that of previous studies.^[8] Thus, the recanalization rate and clinical outcomes in this study were good and comparable to those in large-scale studies, but we lacked sufficient statistical power to determine whether the MPMGW has an effect on outcomes. Despite these limitations, we believe that the results of this study can significantly contribute to the currently available clinical information about rapid recanalization in acute thrombectomy. However, large-scale, multicenter trials must be conducted to validate the results of the current study.



Figure 3: Images of case 4. R indicates the right side. *P* indicates the posterior side. (a) Anterior to posterior view. Right carotid angiographic image showing internal carotid artery occlusion (arrow). (b) Lateral view. Right carotid angiographic image showing internal carotid artery occlusion just distal to the ophthalmic artery (arrow). (c) The tip of the preshaped microguidewire (arrow). (d) Complete recanalization has been achieved

The use of the preshaped MPMGW in acute thrombectomy was effective in terms of both security of procedure and reduction in recanalization time.

Acknowledgment

We would like to thank Seiichiro Watanabe, head of the Medical Innovation Co., Ltd., Tokyo, Japan, for the support and encouragement in carrying out this study.

Financial support and sponsorship

Nil.

Conflicts of interest

There are no conflicts of interest.

References

- 1. Rha JH, Saver JL. The impact of recanalization on ischemic stroke outcome: A meta-analysis. Stroke 2007;38:967-73.
- Sato M, Ohshima T, Ishikawa K, Goto S, Yamamoto T, Izumi T, et al. A novel technique of safe and versatile microguidewire shaping with neuroendovascular therapy: Modified pigtail method. J Neuroendovasc Ther 2017;11:266-71.
- Ohshima T, Ishikawa K, Goto S, Yamamoto T. Relationship between clot quality and microguidewire configuration during endovascular thrombectomy for acute ischemic stroke. World Neurosurg 2017;107:657-62.
- 4. Ohshima T, Imai T, Sato M, Goto S, Yamamoto T, Nishizawa T,

et al. A novel technique for higher success rates of recanalization with stent clot retriever: Corkscrew penetrating method. J Neuroendovasc Ther 2017;11:94-8.

- Ohshima T, Shamim Ul Haq Siddiqi, Miyachi S, Matsuo N, Kawaguchi R, Kato Y, *et al.* Usefulness of modified pigtail-shaped microguidewire guidance for microcatheter navigation in difficult vasculatures during neuroendovascular interventions. Nagoya J Med Sci 2018;80:551-7.
- Ato F, Ohshima T, Miyachi S, Matsuo N, Kawaguchi R, Takayasu M. Efficacy and safety of a modified pigtail-shaped microguidewire during endovascular thrombectomy. Asian J

Neurosurg 2019;14:1040-3.

- Goto S, Ohshima T, Ishikawa K, Yamamoto T, Shimato S, Nishizawa T, et al. A Stent-Retrieving into an Aspiration Catheter with Proximal Balloon (ASAP) technique: A technique of mechanical thrombectomy. World Neurosurg 2018;109:e468-75.
- Goyal M, Menon BK, van Zwam WH, Dippel DW, Mitchell PJ, Demchuk AM, *et al.* Endovascular thrombectomy after large-vessel ischaemic stroke: A meta-analysis of individual patient data from five randomised trials. Lancet 2016;387:1723-31.